

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112621\
 Data File : VX025351.D
 Acq On : 26 Nov 2021 13:16
 Operator : JC/MD
 Sample : M4839-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ESQP4

Quant Time: Nov 26 23:57:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 26 23:56:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	144534	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	130071	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	59995	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	42429	43.015	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.020%
7) Chloroethane-d5	1.666	69	38830	47.697	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.400%
11) 1,1-Dichloroethene-d2	2.306	63	55435	35.247	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.500%
21) 2-Butanone-d5	4.458	46	75950	103.438	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.440%
24) Chloroform-d	5.062	84	85840	49.794	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.580%
26) 1,2-Dichloroethane-d4	5.958	65	55349	51.510	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.020%
32) Benzene-d6	5.976	84	173532	50.339	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.680%
36) 1,2-Dichloropropane-d6	7.312	67	54513	51.075	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.140%
41) Toluene-d8	8.653	98	156505	48.098	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.200%
43) trans-1,3-Dichloroprop...	8.952	79	24659	45.954	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.900%
47) 2-Hexanone-d5	9.384	63	55322	99.575	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.570%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	76906	50.984	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.960%
66) 1,2-Dichlorobenzene-d4	12.323	152	61361	52.146	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.300%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112621\
Data File : VX025351.D
Acq On : 26 Nov 2021 13:16
Operator : JC/MD
Sample : M4839-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ESQP4

Quant Time: Nov 26 23:57:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 26 23:56:11 2021
Response via : Initial Calibration

