

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112621\
 Data File : VX025352.D
 Acq On : 26 Nov 2021 13:42
 Operator : JC/MD
 Sample : M4839-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK002

Quant Time: Nov 26 23:58:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 26 23:56:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	152695	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	138296	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	63942	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	45108	43.287	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.580%
7) Chloroethane-d5	1.660	69	39564	46.001	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.000%
11) 1,1-Dichloroethene-d2	2.306	63	58718	35.339	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.680%
21) 2-Butanone-d5	4.452	46	79486	102.468	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.470%
24) Chloroform-d	5.062	84	88658	48.680	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.360%
26) 1,2-Dichloroethane-d4	5.958	65	57260	50.440	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.880%
32) Benzene-d6	5.976	84	181718	49.578	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.160%
36) 1,2-Dichloropropane-d6	7.312	67	56398	49.698	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.400%
41) Toluene-d8	8.653	98	165525	47.845	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.700%
43) trans-1,3-Dichloroprop...	8.952	79	26380	46.237	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.480%
47) 2-Hexanone-d5	9.384	63	58915	99.735	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.730%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	80191	50.000	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.000%
66) 1,2-Dichlorobenzene-d4	12.323	152	64159	51.158	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112621\
Data File : VX025352.D
Acq On : 26 Nov 2021 13:42
Operator : JC/MD
Sample : M4839-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK002

Quant Time: Nov 26 23:58:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 26 23:56:11 2021
Response via : Initial Calibration

