

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112624\
 Data File : VX044005.D
 Acq On : 26 Nov 2024 10:31
 Operator : JC/MD
 Sample : VX1126MBL01
 Misc : 5.00g/10.0mL/100ul/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1126MBL01

Quant Time: Nov 27 00:47:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	118594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	226731	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	191588	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	70921	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	101446	49.873	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.740%
35) Dibromofluoromethane	5.385	113	78099	48.206	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.420%
50) Toluene-d8	8.641	98	276622	49.532	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.060%
62) 4-Bromofluorobenzene	11.079	95	82744	43.535	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.080%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112624\
Data File : VX044005.D
Acq On : 26 Nov 2024 10:31
Operator : JC/MD
Sample : VX1126MBL01
Misc : 5.00g/10.0mL/100ul/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1126MBL01

Quant Time: Nov 27 00:47:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

