

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112624\
 Data File : VX044010.D
 Acq On : 26 Nov 2024 12:31
 Operator : JC/MD
 Sample : P4846-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-FD-3-20241113

Quant Time: Nov 27 00:50:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	127894	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	241817	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	84284	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	110550	50.397	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.800%
35) Dibromofluoromethane	5.385	113	81222	47.006	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.020%
50) Toluene-d8	8.647	98	290570	48.784	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.560%
62) 4-Bromofluorobenzene	11.079	95	92637	45.700	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.400%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1583	1.571	ug/l #	79
44) Trichloroethene	7.123	130	124313	65.474	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112624\
Data File : VX044010.D
Acq On : 26 Nov 2024 12:31
Operator : JC/MD
Sample : P4846-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-FD-3-20241113

Quant Time: Nov 27 00:50:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

