

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX112624\  
 Data File : VX044017.D  
 Acq On : 26 Nov 2024 15:14  
 Operator : JC/MD  
 Sample : P5003-02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RW-10A-HYD-20241122

Quant Time: Nov 27 00:54:31 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X112124W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Nov 22 03:45:52 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 5.549  | 168   | 98918    | 50.000   | ug/l   | 0.01     |
| 34) 1,4-Difluorobenzene     | 6.757  | 114   | 187238   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 163593   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 68706    | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 84155    | 49.602   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =      | 99.200%  |
| 35) Dibromofluoromethane    | 5.391  | 113   | 60703    | 45.372   | ug/l   | 0.01     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 90.740%  |
| 50) Toluene-d8              | 8.646  | 98    | 229899   | 49.849   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =      | 99.700%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 74787    | 47.648   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 77 - 121 | Recovery | =      | 95.300%  |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 3) Chloromethane            | 1.300  | 50    | 429      | 0.325    | ug/l   | 89       |
| 16) Acetone                 | 2.391  | 43    | 1694     | 2.174    | ug/l   | 95       |
| 60) Dibromochloromethane    | 9.524  | 129   | 937      | 0.728    | ug/l   | 96       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX112624\  
Data File : VX044017.D  
Acq On : 26 Nov 2024 15:14  
Operator : JC/MD  
Sample : P5003-02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
RW-10A-HYD-20241122

Quant Time: Nov 27 00:54:31 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X112124W.M  
Quant Title : SW846 8260  
QLast Update : Fri Nov 22 03:45:52 2024  
Response via : Initial Calibration

