



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112624\
Data File : VX044003.D
Acq On : 26 Nov 2024 08:09
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Title : SW846 8260
Last Update : Fri Nov 22 03:45:52 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	21.8	6446	PASS
75	95	30	60	55.4	16364	PASS
95	95	100	100	100.0	29541	PASS
96	95	5	9	6.2	1841	PASS
173	174	0.00	2	1.3	278	PASS
174	95	50	100	72.2	21339	PASS
175	174	5	9	7.6	1625	PASS
176	174	95	101	98.3	20968	PASS
177	176	5	9	6.8	1416	PASS