

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112625\
 Data File : VX048687.D
 Acq On : 26 Nov 2025 10:01
 Operator : JC/MD
 Sample : VX1126MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1126MBL01

Quant Time: Nov 27 01:33:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	131955	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	271256	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	276843	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	149303	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	111858	60.836	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	121.680%#
35) Dibromofluoromethane	5.367	113	94767	46.159	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.320%
50) Toluene-d8	8.635	98	319390	49.881	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.760%
62) 4-Bromofluorobenzene	11.061	95	136897	57.370	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112625\
Data File : VX048687.D
Acq On : 26 Nov 2025 10:01
Operator : JC/MD
Sample : VX1126MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1126MBL01

Quant Time: Nov 27 01:33:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

