

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112625\
 Data File : VX048687.D
 Acq On : 26 Nov 2025 10:01
 Operator : JC/MD
 Sample : VX1126MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1126MBL01

Quant Time: Nov 27 01:33:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	131955	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	271256	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	276843	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	149303	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	111858	60.836	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	121.680%#
35) Dibromofluoromethane	5.367	113	94767	46.159	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.320%
50) Toluene-d8	8.635	98	319390	49.881	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.760%
62) 4-Bromofluorobenzene	11.061	95	136897	57.370	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.740%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112625\
Data File : VX048687.D
Acq On : 26 Nov 2025 10:01
Operator : JC/MD
Sample : VX1126MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1126MBL01

Quant Time: Nov 27 01:33:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 730
Delta R.T. -0.000 min
Lab File: VX048687.D
Acq: 26 Nov 2025 10:01




Instrument :
MSVOA_X
ClientSampleId :
VX1126MBL01


Tgt Ion:168 Resp: 131955
Ion Ratio Lower Upper
168 100
99 61.6 45.2 67.8



#33
1,2-Dichloroethane-d4
Concen: 60.836 ug/l
RT: 5.934 min Scan# 795
Delta R.T. -0.012 min
Lab File: VX048687.D
Acq: 26 Nov 2025 10:01

Tgt Ion: 65 Resp: 111858
Ion Ratio Lower Upper
65 100
67 51.8 0.0 106.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.745 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX048687.D
Acq: 26 Nov 2025 10:01




Instrument :
MSVOA_X
ClientSampleId :
VX1126MBL01


Tgt Ion:114 Resp: 271256
Ion Ratio Lower Upper
114 100
63 20.0 0.0 41.2
88 16.3 0.0 32.2



#35
Dibromofluoromethane
Concen: 46.159 ug/l
RT: 5.367 min Scan# 702
Delta R.T. -0.006 min
Lab File: VX048687.D
Acq: 26 Nov 2025 10:01

Tgt Ion:113 Resp: 94767
Ion Ratio Lower Upper
113 100
111 103.1 81.9 122.9
192 18.8 15.8 23.6









#50
Toluene-d8
Concen: 49.881 ug/l
RT: 8.635 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX048687.D
Acq: 26 Nov 2025 10:01




Instrument :
MSVOA_X
ClientSampleId :
VX1126MBL01


Tgt Ion: 98 Resp: 319390
Ion Ratio Lower Upper
98 100
100 66.0 53.6 80.4



#62
4-Bromofluorobenzene
Concen: 57.370 ug/l
RT: 11.061 min Scan# 1636
Delta R.T. -0.000 min
Lab File: VX048687.D
Acq: 26 Nov 2025 10:01

Tgt Ion: 95 Resp: 136897
Ion Ratio Lower Upper
95 100
174 83.7 0.0 161.8
176 82.3 0.0 153.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.037 min Scan# 1468
Delta R.T. -0.000 min
Lab File: VX048687.D
Acq: 26 Nov 2025 10:01

Instrument :
MSVOA_X
ClientSampleId :
VX1126MBL01

Tgt Ion:117 Resp: 276843
Ion Ratio Lower Upper
117 100
82 53.0 44.4 66.6
119 32.3 25.1 37.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX048687.D
Acq: 26 Nov 2025 10:01

Tgt Ion:152 Resp: 149303
Ion Ratio Lower Upper
152 100
115 57.3 39.2 117.6
150 157.3 0.0 347.0

