


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112625\
Data File : VX048685.D
Acq On : 26 Nov 2025 08:37
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Title : SW846 8260
Last Update : Sat Nov 08 05:08:11 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1629

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.6	12429	PASS
75	95	30	60	51.4	34299	PASS
95	95	100	100	100.0	66723	PASS
96	95	5	9	6.6	4408	PASS
173	174	0.00	2	1.3	652	PASS
174	95	50	100	77.6	51808	PASS
175	174	5	9	7.5	3901	PASS
176	174	95	101	95.7	49589	PASS
177	176	5	9	6.8	3368	PASS