

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX112719\
 Data File : VX013669.D
 Acq On : 27 Nov 2019 22:51
 Operator : JC/SP
 Sample : VX1127WBS03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1127WBS03

Quant Time: Nov 28 04:19:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	656189	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1008885	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	902149	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	441538	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	362312	47.57	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.14%	
35) Dibromofluoromethane	5.49	113	300545	48.19	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.38%	
50) Toluene-d8	8.71	98	1161383	47.79	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.58%	
62) 4-Bromofluorobenzene	11.14	95	409490	46.73	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	93.46%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.19	85	149540	20.474	ug/l	99
3) Chloromethane	1.32	50	165451	20.189	ug/l	99
4) Vinyl Chloride	1.40	62	194037	21.036	ug/l	98
5) Bromomethane	1.63	94	111249	17.959	ug/l	97
6) Chloroethane	1.72	64	103914	19.797	ug/l	98
7) Trichlorofluoromethane	1.92	101	201560	19.679	ug/l	98
8) Diethyl Ether	2.18	74	95956	20.708	ug/l	98
9) 1,1,2-Trichlorotrifluoroet	2.37	101	122964	19.655	ug/l	98
10) Methyl Iodide	2.50	142	119898	15.843	ug/l	99
11) Tert butyl alcohol	3.03	59	185001	98.400	ug/l	100
12) 1,1-Dichloroethene	2.36	96	125054	20.025	ug/l	99
13) Acrolein	2.28	56	99791	83.464	ug/l	97
14) Allyl chloride	2.72	41	218064	19.242	ug/l	99
15) Acrylonitrile	3.13	53	403177	103.889	ug/l	99
16) Acetone	2.43	43	354936	94.384	ug/l	97
17) Carbon Disulfide	2.56	76	322222	18.100	ug/l	99
18) Methyl Acetate	2.76	43	229840	21.949	ug/l	99
19) Methyl tert-butyl Ether	3.18	73	413115	19.969	ug/l	98
20) Methylene Chloride	2.85	84	144978	20.067	ug/l	94
21) trans-1,2-Dichloroethene	3.15	96	132391	19.461	ug/l	99
22) Diisopropyl ether	3.84	45	438734	20.106	ug/l	94
23) Vinyl Acetate	3.80	43	1804917	99.427	ug/l	99
24) 1,1-Dichloroethane	3.69	63	240901	20.195	ug/l	98
25) 2-Butanone	4.66	43	567371	101.518	ug/l	98
26) 2,2-Dichloropropane	4.58	77	158726	16.082	ug/l	99
27) cis-1,2-Dichloroethene	4.59	96	153864	19.695	ug/l	97
28) Bromochloromethane	5.01	49	90717	19.236	ug/l	100
29) Tetrahydrofuran	5.12	42	365566	106.103	ug/l	99
30) Chloroform	5.20	83	232645	19.230	ug/l	96
31) Cyclohexane	5.58	56	214104	19.029	ug/l	99
32) 1,1,1-Trichloroethane	5.48	97	194698	19.258	ug/l	100
36) 1,1-Dichloropropene	5.79	75	173996	18.999	ug/l	98
37) Ethyl Acetate	4.82	43	217305	21.151	ug/l	100
38) Carbon Tetrachloride	5.77	117	163373	19.346	ug/l	93

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112719\
 Data File : VX013669.D
 Acq On : 27 Nov 2019 22:51
 Operator : JC/SP
 Sample : VX1127WBS03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1127WBS03

Quant Time: Nov 28 04:19:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.46	83	210544	18.599	ug/l	97
40) Benzene	6.14	78	545816	19.568	ug/l	100
41) Methacrylonitrile	5.03	41	116962	20.758	ug/l	99
42) 1,2-Dichloroethane	6.18	62	186328	19.588	ug/l	99
43) Isopropyl Acetate	6.43	43	336061	19.868	ug/l	99
44) Trichloroethene	7.21	130	145611	19.022	ug/l	99
45) 1,2-Dichloropropane	7.51	63	140516	20.135	ug/l	99
46) Dibromomethane	7.65	93	89465	18.875	ug/l	97
47) Bromodichloromethane	7.89	83	173653	19.380	ug/l	96
48) Methyl methacrylate	7.76	41	168593	20.514	ug/l	99
49) 1,4-Dioxane	7.73	88	73897	401.709	ug/l	99
51) 4-Methyl-2-Pentanone	8.64	43	1100049	103.393	ug/l	100
52) Toluene	8.78	92	344010	19.853	ug/l	99
53) t-1,3-Dichloropropene	9.04	75	186835	18.319	ug/l	99
54) cis-1,3-Dichloropropene	8.43	75	209901	18.838	ug/l	97
55) 1,1,2-Trichloroethane	9.21	97	137254	19.656	ug/l	98
56) Ethyl methacrylate	9.17	69	222468	19.816	ug/l	99
57) 1,3-Dichloropropane	9.37	76	236700	19.666	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.31	63	439995	99.960	ug/l	100
59) 2-Hexanone	9.48	43	842175	101.491	ug/l	99
60) Dibromochloromethane	9.58	129	137869	19.025	ug/l	99
61) 1,2-Dibromoethane	9.67	107	146920	19.700	ug/l	100
64) Tetrachloroethene	9.33	164	138143	20.534	ug/l	99
65) Chlorobenzene	10.14	112	371295	19.823	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.22	131	129906	19.314	ug/l	97
67) Ethyl Benzene	10.25	91	641706	19.937	ug/l	100
68) m/p-Xylenes	10.35	106	487700	39.577	ug/l	99
69) o-Xylene	10.70	106	235891	19.628	ug/l	97
70) Styrene	10.71	104	401087	19.839	ug/l	100
71) Bromoform	10.85	173	105663	18.987	ug/l #	100
73) Isopropylbenzene	11.01	105	627647	20.265	ug/l	100
74) N-amyl acetate	10.89	43	284679	20.065	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.26	83	223239	20.784	ug/l	100
76) 1,2,3-Trichloropropane	11.29	75	242198	24.388	ug/l	88
77) Bromobenzene	11.25	156	163946	19.594	ug/l	99
78) n-propylbenzene	11.35	91	694941	20.100	ug/l	98
79) 2-Chlorotoluene	11.42	91	419430	20.223	ug/l	99
80) 1,3,5-Trimethylbenzene	11.50	105	517377	20.382	ug/l	99
81) trans-1,4-Dichloro-2-buten	11.07	75	63695	17.547	ug/l	96
82) 4-Chlorotoluene	11.51	91	480385	19.998	ug/l	99
83) tert-Butylbenzene	11.77	119	516830	20.034	ug/l	100
84) 1,2,4-Trimethylbenzene	11.81	105	527203	20.464	ug/l	99
85) sec-Butylbenzene	11.94	105	596487	20.242	ug/l	100
86) p-Isopropyltoluene	12.06	119	547972	19.989	ug/l	100
87) 1,3-Dichlorobenzene	12.02	146	284311	19.493	ug/l	99
88) 1,4-Dichlorobenzene	12.09	146	284210	19.131	ug/l	99
89) n-Butylbenzene	12.38	91	450221	19.256	ug/l	99
90) Hexachloroethane	12.59	117	94173	19.006	ug/l	98
91) 1,2-Dichlorobenzene	12.38	146	287840	19.974	ug/l	99
92) 1,2-Dibromo-3-Chloropropan	12.99	75	46500	19.880	ug/l	99


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112719\
Data File : VX013669.D
Acq On : 27 Nov 2019 22:51
Operator : JC/SP
Sample : VX1127WBS03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1127WBS03

Quant Time: Nov 28 04:19:16 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	13.64	180	186229	18.542	ug/l	99
94) Hexachlorobutadiene	13.78	225	89150	18.478	ug/l	98
95) Naphthalene	13.83	128	604631	20.215	ug/l	99
96) 1,2,3-Trichlorobenzene	14.01	180	190461	19.153	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 781

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :


VX1127WBS03

Tot Ion:168 Resp: 656189

Ion Ratio Lower Upper




168 100

99 49.4 39.8 59.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#2

Dichlorodifluoromethane

Concen: 20.474 ug/l

RT: 1.19 min Scan# 50

Delta R.T. 0.01 min

Lab File: VX013669.D


Acq: 27 Nov 2019 22:51

Tgt Ion: 85 Resp: 149540

Ion Ratio Lower Upper



85 100


87 30.9 15.7 47.1



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0





#3

Chloromethane

Concen: 20.189 ug/l

RT: 1.32 min Scan# 71

Delta R.T. 0.01 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :





VX1127WBS03

Tot Ion: 50 Resp: 165451

Ion Ratio Lower Upper

50 100

52 33.0 26.0 39.0



#4

Vinyl Chloride

Concen: 21.036 ug/l

RT: 1.40 min Scan# 84

Delta R.T. -0.00 min

Lab File: VX013669.D




Acq: 27 Nov 2019 22:51

Tgt Ion: 62 Resp: 194037

Ion Ratio Lower Upper

62 100

64 31.3 26.0 39.0





#5

Bromomethane

Concen: 17.959 ug/l

RT: 1.63 min Scan# 122

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :


VX1127WBS03

Tot Ion: 94 Resp: 111249

Ion Ratio Lower Upper

94 100

96 93.5 77.0 115.4



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0

1.63

120000

100000

80000



60000

40000

20000

0

Time--> 1.55 1.60 1.65 1.70 1.75



#6

Chloroethane

Concen: 19.797 ug/l

RT: 1.72 min Scan# 136

Delta R.T. 0.01 min

Lab File: VX013669.D


Acq: 27 Nov 2019 22:51

Tgt Ion: 64 Resp: 103914

Ion Ratio Lower Upper

64 100

66 34.8 26.7 40.1



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0

1.72

80000


60000

40000

20000

0

Time--> 1.65 1.70 1.75 1.80









#7

Trichlorofluoromethane
Concen: 19.679 ug/l
RT: 1.92 min Scan# 170
Delta R.T. 0.01 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Instrument :
MSVOA_X
ClientSampleId :
VX1127WBS03




Tot Ion:101 Resp: 201560
Ion Ratio Lower Upper
101 100
103 63.0 51.8 77.6




#8

Diethyl Ether
Concen: 20.708 ug/l
RT: 2.18 min Scan# 212
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Tgt Ion: 74 Resp: 95956
Ion Ratio Lower Upper
74 100
45 96.4 48.9 146.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 19.655 ug/l

RT: 2.37 min Scan# 244

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :

VX1127WBS03


Tot Ion:101 Resp: 122964

Ion Ratio Lower Upper

101 100

85 41.6 33.9 50.9




151 78.7 64.2 96.2



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VX0

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 15.843 ug/l

RT: 2.50 min Scan# 265

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51


Tgt Ion:142 Resp: 119898

Ion Ratio Lower Upper

142 100

127 38.4 31.0 46.4



141 14.1 11.6 17.4




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol
 Concen: 98.400 ug/l
 RT: 3.03 min Scan# 352
 Delta R.T. -0.00 min
 Lab File: VX013669.D
 Acq: 27 Nov 2019 22:51

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1127WBS03




Tot Ion: 59 Resp: 185001
 Ion Ratio Lower Upper
 59 100
 57 9.9 7.8 11.6




#12

1,1-Dichloroethene
 Concen: 20.025 ug/l
 RT: 2.36 min Scan# 242
 Delta R.T. -0.00 min
 Lab File: VX013669.D
 Acq: 27 Nov 2019 22:51

Tgt Ion: 96 Resp: 125054
 Ion Ratio Lower Upper
 96 100
 61 157.9 126.6 190.0
 98 64.2 50.6 76.0





#13

Acrolein

Concen: 83.464 ug/l

RT: 2.28 min Scan# 229

Delta R.T. 0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :





VX1127WBS03

Tot Ion: 56 Resp: 99791

Ion Ratio Lower Upper

56 100

55 72.7 56.2 84.4



#14

Allyl chloride

Concen: 19.242 ug/l

RT: 2.72 min Scan# 300

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51




Tgt Ion: 41 Resp: 218064


Ion Ratio Lower Upper

41 100

39 64.1 51.7 77.5

76 32.0 25.9 38.9





#15

Acrylonitrile

Concen: 103.889 ug/l

RT: 3.13 min Scan# 368

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :


MSVOA_X

ClientSampleId :

VX1127WBS03

Tot Ion: 53 Resp: 403177


Ion	Ratio	Lower	Upper
53	100		
52	82.8	66.2	99.2
51	35.4	29.0	43.6




Abundance Ion 53.00 (52.70 to 53.70): VX0

Ion 52.00 (51.70 to 52.70): VX0

Ion 51.00 (50.70 to 51.70): VX0



Time-->



#16

Acetone

Concen: 94.384 ug/l

RT: 2.43 min Scan# 253


Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51


Tgt Ion: 43 Resp: 354936

Ion	Ratio	Lower	Upper
43	100		
58	29.3	24.6	36.8




Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0







Time-->



#17
Carbon Disulfide
Concen: 18.100 ug/l
RT: 2.56 min Scan# 275
Delta R.T. 0.01 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51




Instrument :
MSVOA_X
ClientSampleId :
VX1127WBS03

Tot Ion: 76 Resp: 322222
Ion Ratio Lower Upper
76 100
78 9.4 7.2 10.8



#18
Methyl Acetate
Concen: 21.949 ug/l
RT: 2.76 min Scan# 308
Delta R.T. 0.01 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Tgt Ion: 43 Resp: 229840
Ion Ratio Lower Upper
43 100
74 24.7 19.5 29.3









#19

Methyl tert-butyl Ether
 Concen: 19.969 ug/l
 RT: 3.18 min Scan# 376
 Delta R.T. -0.00 min
 Lab File: VX013669.D
 Acq: 27 Nov 2019 22:51

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1127WBS03




Tot Ion: 73 Resp: 413115
 Ion Ratio Lower Upper
 73 100
 57 22.0 18.3 27.5




#20

Methylene Chloride
 Concen: 20.067 ug/l
 RT: 2.85 min Scan# 322
 Delta R.T. 0.01 min
 Lab File: VX013669.D
 Acq: 27 Nov 2019 22:51

Tgt Ion: 84 Resp: 144978
 Ion Ratio Lower Upper
 84 100
 49 114.8 99.7 149.5
 51 35.3 29.9 44.9
 86 63.8 52.2 78.4





#21

trans-1,2-Dichloroethene

Concen: 19.461 ug/l

RT: 3.15 min Scan# 372

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :

VX1127WBS03





Tot Ion: 96 Resp: 132391

Ion Ratio Lower Upper

96 100

61 132.8 107.2 160.8

98 63.6 51.4 77.0



#22

Diisopropyl ether

Concen: 20.106 ug/l

RT: 3.84 min Scan# 485

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Tgt Ion: 45 Resp: 438734




Ion Ratio Lower Upper

45 100

43 73.1 52.6 79.0

87 26.3 21.1 31.7

59 12.2 9.2 13.8





#23

Vinyl Acetate

Concen: 99.427 ug/l

RT: 3.80 min Scan# 478

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :





MSVOA_X

ClientSampleId :

VX1127WBS03

Tot Ion: 43 Resp: 1804917

Ion	Ratio	Lower	Upper
43	100		
86	10.5	8.7	13.1



#24

1,1-Dichloroethane

Concen: 20.195 ug/l

RT: 3.69 min Scan# 459




Delta R.T. -0.00 min


Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Tgt Ion: 63 Resp: 240901

Ion	Ratio	Lower	Upper
63	100		
98	7.8	3.5	10.4
100	4.9	2.1	6.3









#25
2-Butanone
Concen: 101.518 ug/l
RT: 4.66 min Scan# 619
Delta R.T. 0.01 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51




Instrument :
MSVOA_X
ClientSampleId :
VX1127WBS03


Tot Ion: 43 Resp: 567371
Ion Ratio Lower Upper
43 100
72 26.9 20.6 31.0



#26
2,2-Dichloropropane
Concen: 16.082 ug/l
RT: 4.58 min Scan# 605
Delta R.T. 0.01 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Tgt Ion: 77 Resp: 158726
Ion Ratio Lower Upper
77 100
97 24.8 12.0 36.1









#27

cis-1,2-Dichloroethene
Concen: 19.695 ug/l
RT: 4.59 min Scan# 607
Delta R.T. 0.01 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Instrument :
MSVOA_X
ClientSampleId :
VX1127WBS03




Tot Ion	Ratio	Lower	Upper
96	100		
61	138.9	0.0	286.4
98	63.2	0.0	127.4




#28

Bromochloromethane
Concen: 19.236 ug/l
RT: 5.01 min Scan# 676
Delta R.T. 0.01 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Tgt Ion	Ratio	Lower	Upper
49	100		
129	3.2	0.0	5.0
130	81.7	65.4	98.0





#29

Tetrahydrofuran

Concen: 106.103 ug/l

RT: 5.12 min Scan# 694

Delta R.T. 0.01 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :

VX1127WBS03


Tot Ion: 42 Resp: 365566

Ion Ratio Lower Upper

42 100

72 45.6 36.3 54.5




71 42.5 33.2 49.8



Abundance Ion 42.00 (41.70 to 42.70): VX0

Ion 72.00 (71.70 to 72.70): VX0

Ion 71.00 (70.70 to 71.70): VX0



#30

Chloroform

Concen: 19.230 ug/l

RT: 5.20 min Scan# 707

Delta R.T. -0.00 min

Lab File: VX013669.D


Acq: 27 Nov 2019 22:51

Tgt Ion: 83 Resp: 232645

Ion Ratio Lower Upper



83 100


85 65.6 50.2 75.4



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0









#31
Cyclohexane
Concen: 19.029 ug/l
RT: 5.58 min Scan# 769
Delta R.T. 0.01 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51




Instrument :
MSVOA_X
ClientSampled :
VX1127WBS03


Tot Ion: 56 Resp: 214104
Ion Ratio Lower Upper
56 100
69 30.8 24.6 36.8
84 86.4 70.0 105.0



#32
1,1,1-Trichloroethane
Concen: 19.258 ug/l
RT: 5.48 min Scan# 754
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Tgt Ion: 97 Resp: 194698
Ion Ratio Lower Upper
97 100
99 64.1 51.6 77.4
61 45.3 36.1 54.1









#33
 1,2-Dichloroethane-d4
 Concen: 47.567 ug/l
 RT: 6.05 min Scan# 847
 Delta R.T. -0.00 min
 Lab File: VX013669.D
 Acq: 27 Nov 2019 22:51




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1127WBS03

Tot Ion: 65 Resp: 362312
 Ion Ratio Lower Upper
 65 100
 67 53.0 0.0 104.4



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.85 min Scan# 978
 Delta R.T. -0.00 min
 Lab File: VX013669.D
 Acq: 27 Nov 2019 22:51

Tgt Ion: 114 Resp: 1008885
 Ion Ratio Lower Upper
 114 100
 63 19.2 0.0 39.6
 88 15.4 0.0 31.4





#35

Dibromofluoromethane

Concen: 48.187 ug/l

RT: 5.49 min Scan# 755

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :

VX1127WBS03


Tot Ion:113 Resp: 300545

Ion Ratio Lower Upper

113 100

111 102.5 83.6 125.4




192 24.3 19.2 28.8



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 18.999 ug/l

RT: 5.79 min Scan# 805

Delta R.T. 0.01 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51


Tgt Ion: 75 Resp: 173996

Ion Ratio Lower Upper

75 100

110 37.4 18.1 54.1



77 31.8 24.5 36.7




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0









#37
Ethyl Acetate
Concen: 21.151 ug/l
RT: 4.82 min Scan# 645
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51




Instrument :
MSVOA_X
ClientSampleId :
VX1127WBS03


Tot Ion:	43	Resp:	217305
Ion	Ratio	Lower	Upper
43	100		
61	13.9	11.0	16.4
70	10.6	8.6	12.8



#38
Carbon Tetrachloride
Concen: 19.346 ug/l
RT: 5.77 min Scan# 801
Delta R.T. -0.01 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Tgt Ion:	117	Resp:	163373
Ion	Ratio	Lower	Upper
117	100		
119	89.5	77.5	116.3
121	29.3	25.0	37.6









#39
Methvlcvclohexane
Concen: 18.599 ug/l
RT: 7.46 min Scan# 1078
Delta R.T. 0.01 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51




Instrument :
MSVOA_X
ClientSampleId :
VX1127WBS03


Tot Ion: 83 Resp: 210544
Ion Ratio Lower Upper
83 100
55 75.0 61.9 92.9
98 47.1 39.4 59.0



#40
Benzene
Concen: 19.568 ug/l
RT: 6.14 min Scan# 861
Delta R.T. 0.01 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Tgt Ion: 78 Resp: 545816
Ion Ratio Lower Upper
78 100
77 23.7 19.0 28.4









#41
Methacrylonitrile
Concen: 20.758 ug/l
RT: 5.03 min Scan# 680
Delta R.T. 0.01 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51




Instrument :
MSVOA_X
ClientSampleId :
VX1127WBS03


Tot Ion:	41	Resp:	116962
Ion	Ratio	Lower	Upper
41	100		
39	55.3	42.9	64.3
67	72.0	57.4	86.0
52	27.9	22.4	33.6



#42
1,2-Dichloroethane
Concen: 19.588 ug/l
RT: 6.18 min Scan# 869
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Tgt Ion:	62	Resp:	186328
Ion	Ratio	Lower	Upper
62	100		
98	10.2	0.0	20.8









#43

Isopropyl Acetate
 Concen: 19.868 ug/l
 RT: 6.43 min Scan# 910
 Delta R.T. -0.00 min
 Lab File: VX013669.D
 Acq: 27 Nov 2019 22:51

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1127WBS03




Tot Ion	Ratio	Lower	Upper
43	100		
61	20.2	16.3	24.5
87	13.2	10.8	16.2




#44

Trichloroethene
 Concen: 19.022 ug/l
 RT: 7.21 min Scan# 1037
 Delta R.T. 0.01 min
 Lab File: VX013669.D
 Acq: 27 Nov 2019 22:51

Tgt Ion	Ratio	Lower	Upper
130	100		
95	94.7	0.0	190.8





#45

1,2-Dichloropropane

Concen: 20.135 ug/l

RT: 7.51 min Scan# 1086

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :





VX1127WBS03

Tot Ion: 63 Resp: 140516

Ion Ratio Lower Upper

63 100

65 31.4 25.5 38.3



#46

Dibromomethane

Concen: 18.875 ug/l

RT: 7.65 min Scan# 1110

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51




Tgt Ion: 93 Resp: 89465


Ion Ratio Lower Upper

93 100

95 86.0 66.2 99.2

174 120.2 93.4 140.0





#47

Bromodichloromethane

Concen: 19.380 ug/l

RT: 7.89 min Scan# 1149

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :

VX1127WBS03


Tot Ion: 83 Resp: 173653

Ion Ratio Lower Upper

83 100

85 61.5 52.1 78.1




127 8.2 7.0 10.4



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0

Ion 126.80 (126.50 to 127.50): V



#48

Methyl methacrylate

Concen: 20.514 ug/l

RT: 7.76 min Scan# 1128

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51


Tgt Ion: 41 Resp: 168593

Ion Ratio Lower Upper

41 100

69 81.4 64.7 97.1



39 55.7 43.8 65.6




Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 39.00 (38.70 to 39.70): VX0





#49

1,4-Dioxane

Concen: 401.709 ug/l

RT: 7.73 min Scan# 1123

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampled :

VX1127WBS03





Tot Ion: 88 Resp: 73897

Ion	Ratio	Lower	Upper
88	100		
43	32.4	25.8	38.6
58	69.6	54.6	82.0

88 100

43 32.4 25.8 38.6

58 69.6 54.6 82.0



#50

Toluene-d8

Concen: 47.793 ug/l

RT: 8.71 min Scan# 1283

Delta R.T. -0.00 min

Lab File: VX013669.D




Acq: 27 Nov 2019 22:51


Tgt Ion: 98 Resp: 1161383

Ion	Ratio	Lower	Upper
98	100		
100	66.0	53.1	79.7

98 100

100 66.0 53.1 79.7





#51

4-Methyl-2-Pentanone

Concen: 103.393 ug/l

RT: 8.64 min Scan# 1271

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :





VX1127WBS03

Tot Ion: 43 Resp: 1100049

Ion Ratio Lower Upper

43 100

58 40.0 32.1 48.1



#52

Toluene

Concen: 19.853 ug/l

RT: 8.78 min Scan# 1294

Delta R.T. -0.01 min

Lab File: VX013669.D




Acq: 27 Nov 2019 22:51


Tgt Ion: 92 Resp: 344010

Ion Ratio Lower Upper

92 100

91 172.2 136.2 204.2









#53
t-1,3-Dichloropropene
Concen: 18.319 ug/l
RT: 9.04 min Scan# 1337
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51




Instrument :
MSVOA_X
ClientSampleId :
VX1127WBS03


Tot Ion: 75 Resp: 186835
Ion Ratio Lower Upper
75 100
77 31.0 24.4 36.6



#54
cis-1,3-Dichloropropene
Concen: 18.838 ug/l
RT: 8.43 min Scan# 1237
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Tgt Ion: 75 Resp: 209901
Ion Ratio Lower Upper
75 100
77 31.4 25.8 38.6
39 52.5 39.9 59.9





#55

1,1,2-Trichloroethane

Concen: 19.656 ug/l

RT: 9.21 min Scan# 1365

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :

VX1127WBS03

Tot Ion: 97 Resp: 137254


Ion Ratio Lower Upper

97 100

83 82.9 67.3 100.9

85 53.9 43.6 65.4

99 62.4 48.2 72.4






Abundance Ion 96.90 (96.60 to 97.60): VX0

Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0

Ion 98.90 (98.60 to 99.60): VX0



#56

Ethyl methacrylate

Concen: 19.816 ug/l

RT: 9.17 min Scan# 1359

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51


Tgt Ion: 69 Resp: 222468

Ion Ratio Lower Upper

69 100

41 67.5 55.0 82.4



39 36.1 28.7 43.1




Abundance Ion 69.00 (68.70 to 69.70): VX0

Ion 41.00 (40.70 to 41.70): VX0

Ion 39.00 (38.70 to 39.70): VX0





#57

1,3-Dichloropropane

Concen: 19.666 ug/l

RT: 9.37 min Scan# 1391

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :





VX1127WBS03

Tot Ion: 76 Resp: 236700

Ion Ratio Lower Upper

76 100

78 32.6 26.0 39.0



#58

2-Chloroethyl Vinyl ether

Concen: 99.960 ug/l

RT: 8.31 min Scan# 1217

Delta R.T. -0.00 min

Lab File: VX013669.D




Acq: 27 Nov 2019 22:51


Tgt Ion: 63 Resp: 439995

Ion Ratio Lower Upper

63 100

106 29.0 23.4 35.0









#59
2-Hexanone
Concen: 101.491 ug/l
RT: 9.48 min Scan# 1410
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51




Instrument :
MSVOA_X
ClientSampleId :
VX1127WBS03


Tot Ion: 43 Resp: 842175
Ion Ratio Lower Upper
43 100
58 55.9 28.1 84.4



#60
Dibromochloromethane
Concen: 19.025 ug/l
RT: 9.58 min Scan# 1426
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Tgt Ion: 129 Resp: 137869
Ion Ratio Lower Upper
129 100
127 78.2 38.5 115.5





#61

1,2-Dibromoethane

Concen: 19.700 ug/l

RT: 9.67 min Scan# 1440

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :


VX1127WBS03

Tot Ion:107 Resp: 146920

Ion Ratio Lower Upper




107 100

109 95.5 76.1 114.1



Abundance Ion 106.90 (106.60 to 107.60): V

Ion 108.80 (108.50 to 109.50): V



#62

4-Bromofluorobenzene

Concen: 46.731 ug/l

RT: 11.14 min Scan# 1681

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51


Tgt Ion: 95 Resp: 409490

Ion Ratio Lower Upper

95 100

174 89.1 0.0 180.0



176 87.1 0.0 173.6




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51




Instrument :
MSVOA_X
ClientSampleId :
VX1127WBS03


Tot Ion:117 Resp: 902149
Ion Ratio Lower Upper
117 100
82 53.2 41.8 62.8
119 33.1 25.8 38.8



#64
Tetrachloroethene
Concen: 20.534 ug/l
RT: 9.33 min Scan# 1385
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Tgt Ion:164 Resp: 138143
Ion Ratio Lower Upper
164 100
166 124.7 99.7 149.5
129 85.2 70.1 105.1
131 85.1 66.9 100.3





#65

Chlorobenzene

Concen: 19.823 ug/l

RT: 10.14 min Scan# 1517

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :


VX1127WBS03

Tot Ion:112 Resp: 371295

Ion Ratio Lower Upper




112 100

114 32.0 26.1 39.1



Abundance Ion 111.90 (111.60 to 112.60): V

Ion 113.90 (113.60 to 114.60): V



#66

1,1,1,2-Tetrachloroethane

Concen: 19.314 ug/l

RT: 10.22 min Scan# 1531

Delta R.T. 0.01 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51


Tgt Ion:131 Resp: 129906

Ion Ratio Lower Upper

131 100

133 98.2 47.7 143.1



119 67.3 32.9 98.6




Abundance Ion 130.80 (130.50 to 131.50): V

Ion 132.80 (132.50 to 133.50): V

Ion 118.80 (118.50 to 119.50): V





#67

Ethyl Benzene

Concen: 19.937 ug/l

RT: 10.25 min Scan# 1535

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :





VX1127WBS03

Tot Ion: 91 Resp: 641706

Ion Ratio Lower Upper

91 100

106 30.4 24.5 36.7



#68

m/p-Xylenes

Concen: 39.577 ug/l

RT: 10.35 min Scan# 1553

Delta R.T. -0.00 min

Lab File: VX013669.D




Acq: 27 Nov 2019 22:51


Tgt Ion: 106 Resp: 487700

Ion Ratio Lower Upper

106 100

91 199.6 159.0 238.4









#69
o-Xylene
Concen: 19.628 ug/l
RT: 10.70 min Scan# 1609
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51




Instrument :
MSVOA_X
ClientSampleId :
VX1127WBS03


Tot Ion:106 Resp: 235891
Ion Ratio Lower Upper
106 100
91 212.6 104.1 312.2



#70
Styrene
Concen: 19.839 ug/l
RT: 10.71 min Scan# 1611
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Tgt Ion:104 Resp: 401087
Ion Ratio Lower Upper
104 100
78 47.9 38.5 57.7
103 54.4 43.7 65.5





#71

Bromoform

Concen: 18.987 ug/l

RT: 10.85 min Scan# 1635

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampled :

VX1127WBS03


Tot Ion:173 Resp: 105663

Ion Ratio Lower Upper

173 100

175 48.8 24.3 72.8




254 0.5 0.2 0.4#



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.07 min Scan# 1835

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51


Tgt Ion:152 Resp: 441538

Ion Ratio Lower Upper

152 100

115 64.3 38.4 115.1



150 162.9 0.0 346.8




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 20.265 ug/l

RT: 11.01 min Scan# 1661

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :





VX1127WBS03

Tot Ion:105 Resp: 627647

Ion Ratio Lower Upper

105 100

120 27.2 13.5 40.5



#74

N-ethyl acetate

Concen: 20.065 ug/l

RT: 10.89 min Scan# 1641

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Tgt Ion: 43 Resp: 284679




Ion Ratio Lower Upper


43 100

70 43.4 34.3 51.5

55 24.9 18.5 27.7

61 25.0 19.8 29.8





#75

1,1,2,2-Tetrachloroethane

Concen: 20.784 ug/l

RT: 11.26 min Scan# 1702

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :

VX1127WBS03


Tot Ion: 83 Resp: 223239

Ion Ratio Lower Upper

83 100

131 10.4 5.1 15.3




85 63.6 31.8 95.3



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): VX0



#76

1,2,3-Trichloropropane

Concen: 24.388 ug/l

RT: 11.29 min Scan# 1707

Delta R.T. -0.00 min

Lab File: VX013669.D


Acq: 27 Nov 2019 22:51

Tgt Ion: 75 Resp: 242198

Ion Ratio Lower Upper



75 100


77 31.2 19.1 57.5



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#77

Bromobenzene

Concen: 19.594 ug/l

RT: 11.25 min Scan# 1700

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :

VX1127WBS03


Tot Ion:156 Resp: 163946

Ion Ratio Lower Upper

156 100

77 150.3 74.4 223.1




158 98.1 48.9 146.6



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VX0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 20.100 ug/l

RT: 11.35 min Scan# 1717

Delta R.T. -0.00 min

Lab File: VX013669.D


Acq: 27 Nov 2019 22:51

Tgt Ion: 91 Resp: 694941

Ion Ratio Lower Upper



91 100

120 24.5 11.8 35.3



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 20.223 ug/l

RT: 11.42 min Scan# 1727

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :


VX1127WBS03

Tot Ion: 91 Resp: 419430

Ion Ratio Lower Upper




91 100

126 34.6 17.5 52.5



Abundance Ion 91.00 (90.70 to 91.70): VX013669.D (-1678) (-)

Ion 125.90 (125.60 to 126.60): VX013669.D (-1678) (-)



#80

1,3,5-Trimethylbenzene

Concen: 20.382 ug/l

RT: 11.50 min Scan# 1741

Delta R.T. -0.00 min

Lab File: VX013669.D


Acq: 27 Nov 2019 22:51

Tgt Ion: 105 Resp: 517377

Ion Ratio Lower Upper



105 100


120 50.0 25.3 75.9



Abundance Ion 105.00 (104.70 to 105.70): VX013669.D (-1692) (-)

Ion 120.00 (119.70 to 120.70): VX013669.D (-1692) (-)





#81

trans-1,4-Dichloro-2-butene

Concen: 17.547 ug/l

RT: 11.07 min Scan# 1670

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :

VX1127WBS03





Tot Ion: 75 Resp: 63695

Ion Ratio Lower Upper

75 100

53 102.1 78.2 117.2

89 45.9 35.4 53.2



#82

4-Chlorotoluene

Concen: 19.998 ug/l

RT: 11.51 min Scan# 1742

Delta R.T. -0.00 min

Lab File: VX013669.D




Acq: 27 Nov 2019 22:51


Tgt Ion: 91 Resp: 480385

Ion Ratio Lower Upper

91 100

126 31.0 15.7 47.0





#83

tert-Butylbenzene

Concen: 20.034 ug/l

RT: 11.77 min Scan# 1785

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :

VX1127WBS03


Tot Ion:119 Resp: 516830

Ion Ratio Lower Upper

119 100

91 54.6 27.2 81.6




134 23.8 11.7 35.1



Abundance Ion 119.00 (118.70 to 119.70): V

Ion 91.00 (90.70 to 91.70): VX0

Ion 134.00 (133.70 to 134.70): V



#84

1,2,4-Trimethylbenzene

Concen: 20.464 ug/l

RT: 11.81 min Scan# 1791

Delta R.T. -0.00 min

Lab File: VX013669.D


Acq: 27 Nov 2019 22:51

Tgt Ion:105 Resp: 527203

Ion Ratio Lower Upper



105 100


120 45.5 23.1 69.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#85

sec-Butylbenzene

Concen: 20.242 ug/l

RT: 11.94 min Scan# 1813

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :


VX1127WBS03

Tot Ion:105 Resp: 596487

Ion Ratio Lower Upper




105 100

134 20.8 10.4 31.2



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 19.989 ug/l

RT: 12.06 min Scan# 1833

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51


Tgt Ion:119 Resp: 547972

Ion Ratio Lower Upper

119 100

134 26.4 13.3 39.9



91 22.5 11.3 33.8




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): VX0





#87

1,3-Dichlorobenzene

Concen: 19.493 ug/l

RT: 12.02 min Scan# 1826





Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :
MSVOA_X
ClientSampleId :
VX1127WBS03

Tot Ion	Ratio	Lower	Upper
146	100		
111	37.6	19.0	57.0
148	64.5	31.8	95.4



#88

1,4-Dichlorobenzene

Concen: 19.131 ug/l




RT: 12.09 min Scan# 1838


Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Tgt Ion	Ratio	Lower	Upper
146	100		
111	38.4	18.9	56.9
148	64.7	31.9	95.5









#89
n-Butylbenzene
Concen: 19.256 ug/l
RT: 12.38 min Scan# 1886
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51




Instrument :
MSVOA_X
ClientSampled :
VX1127WBS03


Tot Ion: 91 Resp: 450221
Ion Ratio Lower Upper
91 100
92 54.0 27.4 82.2
134 27.2 13.5 40.4



#90
Hexachloroethane
Concen: 19.006 ug/l
RT: 12.59 min Scan# 1920
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Tgt Ion: 117 Resp: 94173
Ion Ratio Lower Upper
117 100
201 86.8 44.1 132.4





#91

1,2-Dichlorobenzene

Concen: 19.974 ug/l

RT: 12.38 min Scan# 1886

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51

Instrument :

MSVOA_X

ClientSampleId :

VX1127WBS03





Tot Ion:146 Resp: 287840

Ion Ratio Lower Upper

146 100

111 38.6 19.6 58.7

148 64.4 31.9 95.7



#92

1,2-Dibromo-3-Chloropropane

Concen: 19.880 ug/l

RT: 12.99 min Scan# 1986

Delta R.T. -0.00 min

Lab File: VX013669.D

Acq: 27 Nov 2019 22:51




Tgt Ion: 75 Resp: 46500


Ion Ratio Lower Upper

75 100

155 95.4 47.6 142.8

157 126.6 62.1 186.2









#93
 1,2,4-Trichlorobenzene
 Concen: 18.542 ug/l
 RT: 13.64 min Scan# 2092
 Delta R.T. -0.00 min
 Lab File: VX013669.D
 Acq: 27 Nov 2019 22:51




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1127WBS03


Tot Ion	Ratio	Lower	Upper
180	100		
182	94.0	47.2	141.6
145	28.1	13.8	41.4



#94
 Hexachlorobutadiene
 Concen: 18.478 ug/l
 RT: 13.78 min Scan# 2115
 Delta R.T. 0.01 min
 Lab File: VX013669.D
 Acq: 27 Nov 2019 22:51

Tgt Ion	Ratio	Lower	Upper
225	100		
223	62.5	31.9	95.5
227	62.2	32.1	96.3









#95
Naphthalene
Concen: 20.215 ug/l
RT: 13.83 min Scan# 2123
Delta R.T. -0.00 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Instrument :
MSVOA_X
ClientSampleId :
VX1127WBS03

Tot Ion:128 Resp: 604631
Ion Ratio Lower Upper
128 100
127 12.3 10.3 15.5
129 10.8 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 19.153 ug/l
RT: 14.01 min Scan# 2153
Delta R.T. -0.01 min
Lab File: VX013669.D
Acq: 27 Nov 2019 22:51

Tgt Ion:180 Resp: 190461
Ion Ratio Lower Upper
180 100
182 96.0 47.6 142.8
145 30.6 15.0 45.0

