

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112719\
 Data File : VX013645.D
 Acq On : 27 Nov 2019 12:17
 Operator : JC/SP
 Sample : K5983-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 28 01:03:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	659199	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1052746	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	922102	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	410515	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	379708	49.62	ug/l	0.00
Spiked Amount			Recovery	=	99.24%	
35) Dibromofluoromethane	5.49	113	312422	48.00	ug/l	0.00
Spiked Amount			Recovery	=	96.00%	
50) Toluene-d8	8.71	98	1235643	48.73	ug/l	0.00
Spiked Amount			Recovery	=	97.46%	
62) 4-Bromofluorobenzene	11.13	95	411499	45.00	ug/l	0.00
Spiked Amount			Recovery	=	90.00%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112719\
Data File : VX013645.D
Acq On : 27 Nov 2019 12:17
Operator : JC/SP
Sample : K5983-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Nov 28 01:03:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

