

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112719\  
 Data File : VX013673.D  
 Acq On : 28 Nov 2019 00:24  
 Operator : JC/SP  
 Sample : K6027-07  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 36 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MW-11

Quant Time: Dec 02 06:48:34 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X112619W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Nov 26 15:53:00 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units  | Dev(Min)  |
|-----------------------------|-------|------|----------|-------|--------|-----------|
| 1) Pentafluorobenzene       | 5.65  | 168  | 647019   | 50.00 | ug/l   | 0.00      |
| 34) 1,4-Difluorobenzene     | 6.85  | 114  | 1018091  | 50.00 | ug/l   | 0.00      |
| 63) Chlorobenzene-d5        | 10.11 | 117  | 896402   | 50.00 | ug/l   | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 12.07 | 152  | 393195   | 50.00 | ug/l   | 0.00      |
| System Monitoring Compounds |       |      |          |       |        |           |
| 33) 1,2-Dichloroethane-d4   | 6.05  | 65   | 372920   | 49.65 | ug/l   | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 99.30% |           |
| 35) Dibromofluoromethane    | 5.49  | 113  | 304071   | 48.31 | ug/l   | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 96.62% |           |
| 50) Toluene-d8              | 8.71  | 98   | 1210393  | 49.36 | ug/l   | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 98.72% |           |
| 62) 4-Bromofluorobenzene    | 11.14 | 95   | 398362   | 45.05 | ug/l   | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 90.10% |           |
| Target Compounds            |       |      |          |       |        |           |
| 64) Tetrachloroethene       | 9.33  | 164  | 5054     | 0.756 | ug/l   | Qvalue 90 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112719\  
Data File : VX013673.D  
Acq On : 28 Nov 2019 00:24  
Operator : JC/SP  
Sample : K6027-07  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 36 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
MW-11

Quant Time: Dec 02 06:48:34 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X112619W.M  
Quant Title : SW846 8260  
QLast Update : Tue Nov 26 15:53:00 2019  
Response via : Initial Calibration

