

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112719\
 Data File : VX013678.D
 Acq On : 28 Nov 2019 02:21
 Operator : JC/SP
 Sample : K5968-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GR-OF-20191121

Quant Time: Nov 28 05:41:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	686518	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1093108	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	962462	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	443199	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	393003	49.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.64%	
35) Dibromofluoromethane	5.49	113	322509	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
50) Toluene-d8	8.71	98	1284502	48.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.58%	
62) 4-Bromofluorobenzene	11.14	95	434435	45.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.52%	
Target Compounds						
16) Acetone	2.44	43	15391	3.912	ug/l	99
84) 1,2,4-Trimethylbenzene	11.81	105	10397	0.402	ug/l	97
95) Naphthalene	13.83	128	79357	2.643	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112719\
Data File : VX013678.D
Acq On : 28 Nov 2019 02:21
Operator : JC/SP
Sample : K5968-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GR-OF-20191121

Quant Time: Nov 28 05:41:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

