

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112719\
 Data File : VX013680.D
 Acq On : 28 Nov 2019 03:08
 Operator : JC/SP
 Sample : K5962-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22BR-20191120

Quant Time: Dec 02 07:10:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	703487	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1098352	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	978926	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	440700	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	396202	48.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.04%	
35) Dibromofluoromethane	5.49	113	327149	48.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.36%	
50) Toluene-d8	8.71	98	1290919	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	11.14	95	437530	45.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.72%	
Target Compounds						
16) Acetone	2.43	43	11605	2.879	ug/l	90
27) cis-1,2-Dichloroethene	4.58	96	59432	7.096	ug/l	89
44) Trichloroethene	7.21	130	15150	1.818	ug/l	97
95) Naphthalene	13.83	128	23602	0.791	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112719\
Data File : VX013680.D
Acq On : 28 Nov 2019 03:08
Operator : JC/SP
Sample : K5962-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22BR-20191120

Quant Time: Dec 02 07:10:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

