

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112719\
 Data File : VX013681.D
 Acq On : 28 Nov 2019 03:31
 Operator : JC/SP
 Sample : K6079-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S39-0198

Quant Time: Dec 02 07:13:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	687753	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1074904	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	958723	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	440349	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	384186	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
35) Dibromofluoromethane	5.49	113	320210	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
50) Toluene-d8	8.71	98	1271292	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
62) 4-Bromofluorobenzene	11.14	95	431201	46.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.38%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	7002	0.724	ug/l #	79
16) Acetone	2.43	43	10986	2.787	ug/l #	88
21) trans-1,2-Dichloroethene	3.16	96	13194	1.850	ug/l	99
27) cis-1,2-Dichloroethene	4.59	96	2211440	270.072	ug/l	89
44) Trichloroethene	7.22	130	5873	0.720	ug/l	85
83) tert-Butylbenzene	11.76	119	9932	0.386	ug/l	95
85) sec-Butylbenzene	11.94	105	62169	2.115	ug/l	82
95) Naphthalene	13.83	128	21112	0.708	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112719\
Data File : VX013681.D
Acq On : 28 Nov 2019 03:31
Operator : JC/SP
Sample : K6079-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S39-0198

Quant Time: Dec 02 07:13:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

