

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112719\
 Data File : VX013685.D
 Acq On : 28 Nov 2019 05:04
 Operator : JC/SP
 Sample : K6008-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CALVTANK-1-20191121

Quant Time: Dec 05 10:51:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	661705	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1036185	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	908078	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	409967	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	375891	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
35) Dibromofluoromethane	5.49	113	308281	48.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.26%	
50) Toluene-d8	8.71	98	1223460	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
62) 4-Bromofluorobenzene	11.14	95	406992	45.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.44%	
Target Compounds						
16) Acetone	2.43	43	8521	2.247	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112719\
Data File : VX013685.D
Acq On : 28 Nov 2019 05:04
Operator : JC/SP
Sample : K6008-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CALVTANK-1-20191121

Quant Time: Dec 05 10:51:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

