

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112719\
 Data File : VX013689.D
 Acq On : 28 Nov 2019 06:38
 Operator : JC/SP
 Sample : K5965-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S39-0190

Quant Time: Dec 02 07:47:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	680236	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1071109	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	941098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	424076	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	386447	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
35) Dibromofluoromethane	5.48	113	314048	47.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.86%	
50) Toluene-d8	8.71	98	1258531	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
62) 4-Bromofluorobenzene	11.14	95	422761	45.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.88%	
Target Compounds						
16) Acetone	2.43	43	7883	2.022	ug/l	93
44) Trichloroethene	7.22	130	2769	0.341	ug/l	77
64) Tetrachloroethene	9.33	164	7226	1.030	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112719\
Data File : VX013689.D
Acq On : 28 Nov 2019 06:38
Operator : JC/SP
Sample : K5965-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S39-0190

Quant Time: Dec 02 07:47:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

