

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112719\
 Data File : VX013667.D
 Acq On : 27 Nov 2019 21:41
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 28 04:29:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.557	0.584	-4.8	93	0.00
3 P	Chloromethane	0.624	0.639	-2.4	93	0.00
4 C	Vinyl Chloride	0.703	0.744	-5.8#	94	0.00
5 T	Bromomethane	0.472	0.436	7.6	90	0.00
6 T	Chloroethane	0.400	0.407	-1.7	97	0.00
7 T	Trichlorofluoromethane	0.780	0.793	-1.7	97	0.00
8 T	Diethyl Ether	0.353	0.367	-4.0	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.477	0.479	-0.4	95	0.00
10 T	Methyl Iodide	0.577	0.548	5.0	82	0.00
11 T	Tert butyl alcohol	0.143	0.128	10.5	87	0.00
12 CM	1,1-Dichloroethene	0.476	0.488	-2.5#	95	0.00
13 T	Acrolein	0.088	0.083	5.7	78	0.00
14 T	Allyl chloride	0.864	0.868	-0.5	92	0.00
15 T	Acrylonitrile	0.296	0.304	-2.7	94	0.00
16 T	Acetone	0.287	0.264	8.0	85	0.00
17 T	Carbon Disulfide	1.356	1.290	4.9	90	0.00
18 T	Methyl Acetate	0.798	0.854	-7.0	95	0.00
19 T	Methyl tert-butyl Ether	1.576	1.630	-3.4	95	0.00
20 T	Methylene Chloride	0.551	0.553	-0.4	96	0.00
21 T	trans-1,2-Dichloroethene	0.518	0.522	-0.8	95	0.00
22 T	Diisopropyl ether	1.663	1.733	-4.2	95	0.00
23 T	Vinyl Acetate	1.383	1.442	-4.3	94	0.00
24 P	1,1-Dichloroethane	0.909	0.946	-4.1	96	0.00
25 T	2-Butanone	0.426	0.425	0.2	89	0.00
26 T	2,2-Dichloropropane	0.752	0.685	8.9	85	0.00
27 T	cis-1,2-Dichloroethene	0.595	0.595	0.0	94	0.00
28 T	Bromochloromethane	0.320	0.397	-24.1#	99	0.00
29 T	Tetrahydrofuran	0.263	0.273	-3.8	92	0.00
30 C	Chloroform	0.922	0.919	0.3#	94	0.00
31 T	Cyclohexane	0.857	0.851	0.7	93	0.00
32 T	1,1,1-Trichloroethane	0.770	0.775	-0.6	94	0.00
33 S	1,2-Dichloroethane-d4	0.580	0.572	1.4	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.309	0.300	2.9	95	0.00
36 T	1,1-Dichloropropene	0.454	0.449	1.1	95	0.00
37 T	Ethyl Acetate	0.509	0.520	-2.2	93	0.00
38 T	Carbon Tetrachloride	0.419	0.421	-0.5	94	0.00
39 T	Methylcyclohexane	0.561	0.535	4.6	92	0.00
40 TM	Benzene	1.382	1.369	0.9	95	0.00
41 T	Methacrylonitrile	0.279	0.289	-3.6	94	0.00
42 TM	1,2-Dichloroethane	0.471	0.471	0.0	96	0.00
43 T	Isopropyl Acetate	0.838	0.843	-0.6	94	0.00
44 TM	Trichloroethene	0.379	0.364	4.0	93	0.00
45 C	1,2-Dichloropropane	0.346	0.351	-1.4#	96	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX112719\
 Data File : VX013667.D
 Acq On : 27 Nov 2019 21:41
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 28 04:29:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.235	0.230	2.1	95	0.00
47 T	Bromodichloromethane	0.444	0.455	-2.5	96	0.00
48 T	Methyl methacrylate	0.407	0.419	-2.9	93	0.00
49 T	1,4-Dioxane	0.009	0.008	11.1	85	0.00
50 S	Toluene-d8	1.204	1.165	3.2	95	0.00
51 T	4-Methyl-2-Pentanone	0.527	0.534	-1.3	94	0.00
52 CM	Toluene	0.859	0.854	0.6#	95	0.00
53 T	t-1,3-Dichloropropene	0.505	0.498	1.4	91	0.00
54 T	cis-1,3-Dichloropropene	0.552	0.555	-0.5	92	0.00
55 T	1,1,2-Trichloroethane	0.346	0.349	-0.9	96	0.00
56 T	Ethyl methacrylate	0.556	0.572	-2.9	96	0.00
57 T	1,3-Dichloropropane	0.597	0.593	0.7	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.218	0.218	0.0	94	0.00
59 T	2-Hexanone	0.411	0.405	1.5	91	0.00
60 T	Dibromochloromethane	0.359	0.369	-2.8	94	0.00
61 T	1,2-Dibromoethane	0.370	0.366	1.1	95	0.00
62 S	4-Bromofluorobenzene	0.434	0.419	3.5	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.373	0.372	0.3	98	0.00
65 PM	Chlorobenzene	1.038	1.022	1.5	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.380	-1.9	95	0.00
67 C	Ethyl Benzene	1.784	1.800	-0.9#	94	0.00
68 T	m/p-Xylenes	0.683	0.687	-0.6	95	0.00
69 T	o-Xylene	0.666	0.671	-0.8	94	0.00
70 T	Styrene	1.120	1.155	-3.1	95	0.00
71 P	Bromoform	0.308	0.319	-3.6	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.507	3.547	-1.1	94	0.00
74 T	N-amyl acetate	1.607	1.678	-4.4	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.216	1.248	-2.6	95	0.00
76 T	1,2,3-Trichloropropane	1.125	0.979	13.0	79	0.00
77 T	Bromobenzene	0.948	0.938	1.1	94	0.00
78 T	n-propylbenzene	3.915	4.022	-2.7	94	0.00
79 T	2-Chlorotoluene	2.349	2.377	-1.2	95	0.00
80 T	1,3,5-Trimethylbenzene	2.875	2.978	-3.6	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.411	0.404	1.7	89	0.00
82 T	4-Chlorotoluene	2.720	2.737	-0.6	94	0.00
83 T	tert-Butylbenzene	2.921	2.994	-2.5	95	0.00
84 T	1,2,4-Trimethylbenzene	2.917	2.990	-2.5	95	0.00
85 T	sec-Butylbenzene	3.337	3.421	-2.5	94	0.00
86 T	p-Isopropyltoluene	3.104	3.131	-0.9	93	0.00
87 T	1,3-Dichlorobenzene	1.652	1.638	0.8	94	0.00
88 T	1,4-Dichlorobenzene	1.682	1.623	3.5	94	0.00
89 T	n-Butylbenzene	2.648	2.707	-2.2	93	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112719\
Data File : VX013667.D
Acq On : 27 Nov 2019 21:41
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 28 04:29:36 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.561	0.559	0.4	94	0.00
91 T	1,2-Dichlorobenzene	1.632	1.631	0.1	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.265	0.269	-1.5	92	0.00
93 T	1,2,4-Trichlorobenzene	1.137	1.103	3.0	91	0.00
94 T	Hexachlorobutadiene	0.546	0.518	5.1	95	0.00
95 T	Naphthalene	3.387	3.547	-4.7	93	0.00
96 T	1,2,3-Trichlorobenzene	1.126	1.141	-1.3	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6