

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112719\
 Data File : VX013646.D
 Acq On : 27 Nov 2019 12:40
 Operator : JC/SP
 Sample : K5983-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Nov 28 01:06:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	666130	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1063170	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	939455	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	418129	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	381260	49.31	ug/l	0.00
Spiked Amount			Recovery	=	98.62%	
35) Dibromofluoromethane	5.48	113	316795	48.20	ug/l	0.00
Spiked Amount			Recovery	=	96.40%	
50) Toluene-d8	8.71	98	1246758	48.69	ug/l	0.00
Spiked Amount			Recovery	=	97.38%	
62) 4-Bromofluorobenzene	11.14	95	415597	45.01	ug/l	0.00
Spiked Amount			Recovery	=	90.02%	


Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112719\
Data File : VX013646.D
Acq On : 27 Nov 2019 12:40
Operator : JC/SP
Sample : K5983-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Nov 28 01:06:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 782

Delta R.T. 0.01 min

Lab File: VX013646.D

Acq: 27 Nov 2019 12:40

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-WATER-REF-4

Tot Ion:168 Resp: 666130

Ion Ratio Lower Upper

168 100

99 47.5 39.8 59.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.65

250000

200000

150000



100000

50000

0

Time-->

5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 49.308 ug/l

RT: 6.05 min Scan# 847

Delta R.T. -0.00 min

Lab File: VX013646.D


Acq: 27 Nov 2019 12:40

Tgt Ion: 65 Resp: 381260

Ion Ratio Lower Upper

65 100

67 52.8 0.0 104.4



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.05

150000


100000


50000

0

Time-->

6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 978

Delta R.T. -0.00 min

Lab File: VX013646.D

Acq: 27 Nov 2019 12:40

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-WATER-REF-4


Tot Ion:114 Resp: 1063170

Ion Ratio Lower Upper

114 100

63 20.2 0.0 39.6




88 15.5 0.0 31.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 48.199 ug/l

RT: 5.48 min Scan# 754

Delta R.T. -0.01 min

Lab File: VX013646.D

Acq: 27 Nov 2019 12:40


Tgt Ion:113 Resp: 316795

Ion Ratio Lower Upper

113 100

111 101.9 83.6 125.4



192 23.8 19.2 28.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.687 ug/l

RT: 8.71 min Scan# 1283

Delta R.T. -0.00 min

Lab File: VX013646.D

Acq: 27 Nov 2019 12:40

Instrument :

MSVOA_X

ClientSampleId :





STORAGE-BLANK-WATER-REF-4

Tot Ion: 98 Resp: 1246758

Ion Ratio Lower Upper

98 100

100 66.0 53.1 79.7



#62

4-Bromofluorobenzene

Concen: 45.006 ug/l

RT: 11.14 min Scan# 1681

Delta R.T. -0.00 min

Lab File: VX013646.D

Acq: 27 Nov 2019 12:40




Tgt Ion: 95 Resp: 415597


Ion Ratio Lower Upper

95 100

174 91.0 0.0 180.0

176 87.5 0.0 173.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.11 min Scan# 1513

Delta R.T. -0.00 min

Lab File: VX013646.D

Acq: 27 Nov 2019 12:40

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-WATER-REF-4


Tot Ion:117 Resp: 939455

Ion Ratio Lower Upper

117 100

82 52.5 41.8 62.8




119 31.8 25.8 38.8



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.07 min Scan# 1835

Delta R.T. -0.00 min

Lab File: VX013646.D

Acq: 27 Nov 2019 12:40


Tgt Ion:152 Resp: 418129

Ion Ratio Lower Upper

152 100

115 55.0 38.4 115.1

150 154.9 0.0 346.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

