

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112719\
 Data File : VX013657.D
 Acq On : 27 Nov 2019 17:24
 Operator : JC/SP
 Sample : K5962-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-20191120

Quant Time: Nov 28 03:36:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	649477	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1047249	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	941830	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	441097	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	377788	50.11	ug/l	0.00
Spiked Amount						
			Recovery	=	100.22%	
35) Dibromofluoromethane	5.49	113	312209	48.22	ug/l	0.00
Spiked Amount						
			Recovery	=	96.44%	
50) Toluene-d8	8.71	98	1238448	49.10	ug/l	0.00
Spiked Amount						
			Recovery	=	98.20%	
62) 4-Bromofluorobenzene	11.14	95	424346	46.65	ug/l	0.00
Spiked Amount						
			Recovery	=	93.30%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112719\
Data File : VX013657.D
Acq On : 27 Nov 2019 17:24
Operator : JC/SP
Sample : K5962-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-20191120

Quant Time: Nov 28 03:36:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 781

Delta R.T. 0.00 min

Lab File: VX013657.D

Acq: 27 Nov 2019 17:24

Instrument :

MSVOA_X

ClientSampleId :


TB-01-20191120

Tot Ion:168 Resp: 649477

Ion Ratio Lower Upper

168 100

99 51.6 39.8 59.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.65

250000

200000



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 50.111 ug/l

RT: 6.05 min Scan# 847

Delta R.T. 0.00 min

Lab File: VX013657.D


Acq: 27 Nov 2019 17:24

Tgt Ion: 65 Resp: 377788

Ion Ratio Lower Upper

65 100

67 52.8 0.0 104.4



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.05


150000


100000

50000

0

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 978

Delta R.T. 0.00 min

Lab File: VX013657.D

Acq: 27 Nov 2019 17:24

Instrument :

MSVOA_X

ClientSampleId :

TB-01-20191120


Tot Ion:114 Resp: 1047249

Ion Ratio Lower Upper

114 100

63 19.8 0.0 39.6




88 15.6 0.0 31.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 48.223 ug/l

RT: 5.49 min Scan# 755

Delta R.T. 0.00 min

Lab File: VX013657.D

Acq: 27 Nov 2019 17:24


Tgt Ion:113 Resp: 312209

Ion Ratio Lower Upper

113 100

111 103.0 83.6 125.4



192 23.6 19.2 28.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.097 ug/l

RT: 8.71 min Scan# 1283

Delta R.T. 0.00 min

Lab File: VX013657.D

Acq: 27 Nov 2019 17:24

Instrument :

MSVOA_X

ClientSampleId :


TB-01-20191120

Tot Ion: 98 Resp: 1238448

Ion Ratio Lower Upper




98 100

100 66.3 53.1 79.7



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V



#62

4-Bromofluorobenzene

Concen: 46.652 ug/l

RT: 11.14 min Scan# 1681

Delta R.T. 0.00 min

Lab File: VX013657.D

Acq: 27 Nov 2019 17:24


Tgt Ion: 95 Resp: 424346

Ion Ratio Lower Upper

95 100

174 91.1 0.0 180.0



176 87.8 0.0 173.6




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. 0.00 min
Lab File: VX013657.D
Acq: 27 Nov 2019 17:24

Instrument :
MSVOA_X
ClientSampleId :
TB-01-20191120

Tot Ion:117 Resp: 941830
Ion Ratio Lower Upper
117 100
82 52.5 41.8 62.8
119 32.7 25.8 38.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. 0.00 min
Lab File: VX013657.D
Acq: 27 Nov 2019 17:24

Tgt Ion:152 Resp: 441097
Ion Ratio Lower Upper
152 100
115 56.0 38.4 115.1
150 158.7 0.0 346.8

