

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX112719\
 Data File : VX013665.D
 Acq On : 27 Nov 2019 20:31
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

apatel
 12/2/2019 3:18:17 PM

Quant Time: Nov 28 04:25:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	644222	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1001938	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	905806	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	453633	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.05	65	364255	48.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.42%	
35) Dibromofluoromethane	5.49	113	304188	49.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.22%	
50) Toluene-d8	8.71	98	1172486	48.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.16%	
62) 4-Bromofluorobenzene	11.14	95	419749	48.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.46%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.19	85	369327	51.506	ug/l	98
3) Chloromethane	1.32	50	400805	49.815	ug/l	97
4) Vinyl Chloride	1.40	62	472631	52.191	ug/l	97
5) Bromomethane	1.63	94	229741	37.776	ug/l	99
6) Chloroethane	1.71	64	258184	50.102	ug/l	97
7) Trichlorofluoromethane	1.92	101	496509	49.376	ug/l	98
8) Diethyl Ether	2.18	74	230785	50.730	ug/l	97
9) 1,1,2-Trichlorotrifluoroet	2.37	101	302908	49.317	ug/l	99
10) Methyl Iodide	2.50	142	297991	40.107	ug/l	100
11) Tert butyl alcohol	3.03	59	435948	236.183	ug/l	98
12) 1,1-Dichloroethene	2.36	96	308929	50.389	ug/l	100
13) Acrolein	2.28	56	265307	215.767	ug/l	99
14) Allyl chloride	2.72	41	556080	49.980	ug/l	100
15) Acrylonitrile	3.13	53	974894	255.873	ug/l	99
16) Acetone	2.44	43	846363	229.245	ug/l	100
17) Carbon Disulfide	2.56	76	820904	46.970	ug/l	100
18) Methyl Acetate	2.76	43	546361	53.144	ug/l	100
19) Methyl tert-butyl Ether	3.19	73	1029594	50.694	ug/l	100
20) Methylene Chloride	2.84	84	349514	49.277	ug/l	98
21) trans-1,2-Dichloroethene	3.16	96	331618	49.652	ug/l	99
22) Diisopropyl ether	3.84	45	1097945	51.250	ug/l	96
23) Vinyl Acetate	3.80	43	4611864	258.772	ug/l	99
24) 1,1-Dichloroethane	3.69	63	600416	51.267	ug/l	99
25) 2-Butanone	4.66	43	1389477	253.232	ug/l	98
26) 2,2-Dichloropropane	4.58	77	436446	45.042	ug/l	100
27) cis-1,2-Dichloroethene	4.59	96	377461	49.212	ug/l	98
28) Bromochloromethane	5.00	49	248725	53.743	ug/l	100
29) Tetrahydrofuran	5.11	42	900447	266.204	ug/l	99
30) Chloroform	5.20	83	575873	48.485	ug/l	97
31) Cyclohexane	5.57	56	541387	49.011	ug/l	96
32) 1,1,1-Trichloroethane	5.48	97	498988	50.274	ug/l	99
36) 1,1-Dichloropropene	5.79	75	437757	48.131	ug/l	98
37) Ethyl Acetate	4.82	43	528403	51.788	ug/l	99
38) Carbon Tetrachloride	5.78	117	418630	49.917	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX112719\
 Data File : VX013665.D
 Acq On : 27 Nov 2019 20:31
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleID :
 VSTDCCC050EC

Manual Integrations
 APPROVED

apatel
 12/2/2019 3:18:17 PM

Quant Time: Nov 28 04:25:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.46	83	538828	47.928	ug/l	98
40) Benzene	6.14	78	1361184	49.139	ug/l	99
41) Methacrylonitrile	5.03	41	292281	52.233	ug/l	99
42) 1,2-Dichloroethane	6.19	62	471424	49.902	ug/l	99
43) Isopropyl Acetate	6.43	43	842250	50.140	ug/l	100
44) Trichloroethene	7.21	130	365188	48.036	ug/l	93
45) 1,2-Dichloropropane	7.51	63	354199	51.106	ug/l	98
46) Dibromomethane	7.65	93	230118	48.887	ug/l	100
47) Bromodichloromethane	7.89	83	446184	50.141	ug/l	100
48) Methyl methacrylate	7.76	41	424964	52.068	ug/l	98
49) 1,4-Dioxane	7.73	88	183982	1007.073	ug/l	97
51) 4-Methyl-2-Pentanone	8.64	43	2742736	259.575	ug/l	100
52) Toluene	8.78	92	849919	49.389	ug/l	98
53) t-1,3-Dichloropropene	9.04	75	497386	49.106	ug/l	96
54) cis-1,3-Dichloropropene	8.43	75	555619	50.210	ug/l	98
55) 1,1,2-Trichloroethane	9.21	97	351005	50.615	ug/l	99
56) Ethyl methacrylate	9.17	69	573766	51.462	ug/l	99
57) 1,3-Dichloropropane	9.37	76	588117	49.201	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.31	63	1115801	255.251	ug/l	100
59) 2-Hexanone	9.48	43	2099907	254.816	ug/l	99
60) Dibromochloromethane	9.58	129	363143	50.458	ug/l	99
61) 1,2-Dibromoethane	9.67	107	365095	49.293	ug/l	98
64) Tetrachloroethene	9.33	164	320242	47.408	ug/l	97
65) Chlorobenzene	10.14	112	911934	48.491	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.21	131	339952	50.339	ug/l	100
67) Ethyl Benzene	10.25	91	1626246	50.322	ug/l	99
68) m/p-Xylenes	10.36	106	1234393	99.768	ug/l	99
69) o-Xylene	10.70	106	600529	49.766	ug/l	100
70) Styrene	10.71	104	1026435	50.567	ug/l	100
71) Bromoform	10.85	173	279254	49.976	ug/l	99
73) Isopropylbenzene	11.01	105	1572029	49.403	ug/l	99
74) N-amyl acetate	10.89	43	750344	51.476	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.26	83	559401	50.692	ug/l	99
76) 1,2,3-Trichloropropane	11.29	75	439652m	43.090	ug/l	
77) Bromobenzene	11.25	156	414373	48.203	ug/l	98
78) n-propylbenzene	11.35	91	1771102	49.860	ug/l	99
79) 2-Chlorotoluene	11.42	91	1053226	49.428	ug/l	100
80) 1,3,5-Trimethylbenzene	11.50	105	1320579	50.636	ug/l	100
81) trans-1,4-Dichloro-2-buten	11.07	75	180775	48.473	ug/l	99
82) 4-Chlorotoluene	11.51	91	1215353	49.245	ug/l	99
83) tert-Butylbenzene	11.76	119	1336236	50.415	ug/l	99
84) 1,2,4-Trimethylbenzene	11.81	105	1328589	50.195	ug/l	99
85) sec-Butylbenzene	11.94	105	1524781	50.366	ug/l	99
86) p-Isopropyltoluene	12.06	119	1410659	50.087	ug/l	99
87) 1,3-Dichlorobenzene	12.02	146	716165	47.792	ug/l	99
88) 1,4-Dichlorobenzene	12.09	146	712092	46.654	ug/l	99
89) n-Butylbenzene	12.39	91	1201867	50.034	ug/l	100
90) Hexachloroethane	12.59	117	250832	49.272	ug/l	99
91) 1,2-Dichlorobenzene	12.39	146	721511	48.732	ug/l	99
92) 1,2-Dibromo-3-Chloropropan	12.99	75	122511	50.981	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112719\
Data File : VX013665.D
Acq On : 27 Nov 2019 20:31
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Manual Integrations
APPROVED

apatel
12/2/2019 3:18:17 PM

Quant Time: Nov 28 04:25:41 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	13.64	180	487304	47.226	ug/l	98
94) Hexachlorobutadiene	13.78	225	225693	45.532	ug/l	100
95) Naphthalene	13.83	128	1599758	52.060	ug/l	100
96) 1,2,3-Trichlorobenzene	14.01	180	500800	49.018	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 782

Delta R.T. 0.01 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tot Ion:168 Resp: 644222

Ion Ratio Lower Upper

168 100


99 47.6 39.8 59.8

Manual Integrations

APPROVED




apatel

12/2/2019 3:18:17 PM



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#2

Dichlorodifluoromethane

Concen: 51.506 ug/l

RT: 1.19 min Scan# 50

Delta R.T. 0.01 min

Lab File: VX013665.D


Acq: 27 Nov 2019 20:31

Tgt Ion: 85 Resp: 369327

Ion Ratio Lower Upper



85 100


87 32.7 15.7 47.1



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0





#3

Chloromethane

Concen: 49.815 ug/l

RT: 1.32 min Scan# 71

Delta R.T. 0.01 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tot Ion: 50 Resp: 400805

Ion Ratio Lower Upper





50 100

52 33.9 26.0 39.0

Manual Integrations
APPROVED

apatel

12/2/2019 3:18:17 PM



#4

Vinyl Chloride

Concen: 52.191 ug/l

RT: 1.40 min Scan# 84

Delta R.T. 0.00 min

Lab File: VX013665.D




Acq: 27 Nov 2019 20:31

Tgt Ion: 62 Resp: 472631

Ion Ratio Lower Upper

62 100

64 30.6 26.0 39.0





#5

Bromomethane

Concen: 37.776 ug/l

RT: 1.63 min Scan# 122

Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tot Ion: 94 Resp: 229741

Ion Ratio Lower Upper





94 100

96 95.6 77.0 115.4

Manual Integrations
APPROVED

apatel

12/2/2019 3:18:17 PM



#6

Chloroethane

Concen: 50.102 ug/l

RT: 1.71 min Scan# 135

Delta R.T. 0.00 min

Lab File: VX013665.D




Acq: 27 Nov 2019 20:31

Tgt Ion: 64 Resp: 258184

Ion Ratio Lower Upper

64 100

66 31.5 26.7 40.1





#7





Trichlorofluoromethane
Concen: 49.376 ug/l
RT: 1.92 min Scan# 169
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
101	100		
103	66.0	51.8	77.6

Manual Integrations
APPROVED




apatel
12/2/2019 3:18:17 PM




#8

Diethyl Ether
Concen: 50.730 ug/l
RT: 2.18 min Scan# 212
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
74	100		
45	95.1	48.9	146.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 49.317 ug/l

RT: 2.37 min Scan# 244

Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050EC

Tot Ion:	101	Resp:	302908
Ion Ratio	Lower	Upper	
101	100		
85	41.2	33.9	50.9
151	81.1	64.2	96.2

Manual Integrations
APPROVED

apatel
12/2/2019 3:18:17 PM



#10

Methyl Iodide

Concen: 40.107 ug/l




RT: 2.50 min Scan# 265


Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Tgt Ion:	142	Resp:	297991
Ion Ratio	Lower	Upper	
142	100		
127	38.8	31.0	46.4
141	14.5	11.6	17.4





#11





Tert butyl alcohol
Concen: 236.183 ug/l
RT: 3.03 min Scan# 352
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
59	100		
57	10.5	7.8	11.6

Manual Integrations
APPROVED




apatel
12/2/2019 3:18:17 PM




#12

1,1-Dichloroethene
Concen: 50.389 ug/l
RT: 2.36 min Scan# 242
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
96	100		
61	158.0	126.6	190.0
98	63.7	50.6	76.0









#13
Acrolein
Concen: 215.767 ug/l
RT: 2.28 min Scan# 229
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
56	100		
55	69.8	56.2	84.4




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#14
Allyl chloride
Concen: 49.980 ug/l
RT: 2.72 min Scan# 300
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
41	100		
39	64.7	51.7	77.5
76	31.7	25.9	38.9





#15

Acrylonitrile

Concen: 255.873 ug/l

RT: 3.13 min Scan# 368

Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050EC

Tot Ion:	53	Resp:	974894
Ion Ratio		Lower	Upper
53	100		
52	83.2	66.2	99.2
51	35.9	29.0	43.6

Manual Integrations
APPROVED

apatel
12/2/2019 3:18:17 PM



#16

Acetone

Concen: 229.245 ug/l




RT: 2.44 min Scan# 254


Delta R.T. 0.01 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Tgt Ion:	43	Resp:	846363
Ion Ratio		Lower	Upper
43	100		
58	30.4	24.6	36.8









#17
Carbon Disulfide
Concen: 46.970 ug/l
RT: 2.56 min Scan# 274
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
76	100		
78	9.1	7.2	10.8




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#18
Methyl Acetate
Concen: 53.144 ug/l
RT: 2.76 min Scan# 308
Delta R.T. 0.01 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
43	100		
74	24.4	19.5	29.3









#19
Methvl tert-butvl Ether
Concen: 50.694 ug/l
RT: 3.19 min Scan# 377
Delta R.T. 0.01 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion: 73 Resp: 1029594
Ion Ratio Lower Upper
73 100
57 22.7 18.3 27.5




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#20
Methylene Chloride
Concen: 49.277 ug/l
RT: 2.84 min Scan# 321
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion: 84 Resp: 349514
Ion Ratio Lower Upper
84 100
49 121.4 99.7 149.5
51 37.5 29.9 44.9
86 63.8 52.2 78.4





#21

trans-1,2-Dichloroethene

Concen: 49.652 ug/l

RT: 3.16 min Scan# 372

Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050EC

Tot Ion:	96	Resp:	331618
Ion	Ratio	Lower	Upper
96	100		
61	135.8	107.2	160.8
98	63.4	51.4	77.0

Manual Integrations
APPROVED

apatel
12/2/2019 3:18:17 PM



#22

Diisopropyl ether

Concen: 51.250 ug/l




RT: 3.84 min Scan# 485


Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Tgt Ion:	45	Resp:	1097945
Ion	Ratio	Lower	Upper
45	100		
43	70.1	52.6	79.0
87	27.6	21.1	31.7
59	11.4	9.2	13.8









#23
 Vinyl Acetate
 Concen: 258.772 ug/l
 RT: 3.80 min Scan# 478
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion: 43 Resp: 4611864
 Ion Ratio Lower Upper
 43 100
 86 10.6 8.7 13.1




Manual Integrations
 APPROVED


apatel
 12/2/2019 3:18:17 PM



#24
 1,1-Dichloroethane
 Concen: 51.267 ug/l
 RT: 3.69 min Scan# 459
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Tgt Ion: 63 Resp: 600416
 Ion Ratio Lower Upper
 63 100
 98 6.9 3.5 10.4
 100 4.6 2.1 6.3





#25

2-Butanone

Concen: 253.232 ug/l

RT: 4.66 min Scan# 619

Delta R.T. 0.01 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tot Ion: 43 Resp: 1389477

Ion Ratio Lower Upper

43 100





72 26.7 20.6 31.0

Manual Integrations

APPROVED

apatel

12/2/2019 3:18:17 PM



#26

2,2-Dichloropropane

Concen: 45.042 ug/l

RT: 4.58 min Scan# 605

Delta R.T. 0.01 min

Lab File: VX013665.D




Acq: 27 Nov 2019 20:31


Tgt Ion: 77 Resp: 436446

Ion Ratio Lower Upper

77 100

97 24.1 12.0 36.1





#27





cis-1,2-Dichloroethene
Concen: 49.212 ug/l
RT: 4.59 min Scan# 607
Delta R.T. 0.01 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Resp	Lower	Upper
96	100	377461		
61	141.0	0.0	286.4	
98	65.5	0.0	127.4	

Manual Integrations
APPROVED




apatel
12/2/2019 3:18:17 PM




#28

Bromochloromethane
Concen: 53.743 ug/l
RT: 5.00 min Scan# 675
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Resp	Lower	Upper
49	100	248725		
129	2.4	0.0	5.0	
130	81.6	65.4	98.0	









#29
Tetrahydrofuran
Concen: 266.204 ug/l
RT: 5.11 min Scan# 693
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
42	100		
72	45.4	36.3	54.5
71	42.1	33.2	49.8




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#30
Chloroform
Concen: 48.485 ug/l
RT: 5.20 min Scan# 708
Delta R.T. 0.01 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
83	100		
85	64.8	50.2	75.4









#31
Cyclohexane
Concen: 49.011 ug/l
RT: 5.57 min Scan# 768
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
56	100		
69	29.3	24.6	36.8
84	83.0	70.0	105.0




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#32
1,1,1-Trichloroethane
Concen: 50.274 ug/l
RT: 5.48 min Scan# 754
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
97	100		
99	64.0	51.6	77.4
61	44.7	36.1	54.1





#33

1,2-Dichloroethane-d4

Concen: 48.711 ug/l

RT: 6.05 min Scan# 847

Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tot Ion: 65 Resp: 364255

Ion Ratio Lower Upper

65 100





67 53.4 0.0 104.4

Manual Integrations

APPROVED

apatel

12/2/2019 3:18:17 PM



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 978

Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31




Tgt Ion: 114 Resp: 1001938

Ion Ratio Lower Upper

114 100

63 20.5 0.0 39.6

88 15.8 0.0 31.4









#35
 Dibromofluoromethane
 Concen: 49.109 ug/l
 RT: 5.49 min Scan# 755
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion	Ratio	Resp	Lower	Upper
113	100	304188		
111	102.4	83.6	125.4	
192	23.6	19.2	28.8	




Manual Integrations
 APPROVED


apatel
 12/2/2019 3:18:17 PM



#36
 1,1-Dichloropropene
 Concen: 48.131 ug/l
 RT: 5.79 min Scan# 805
 Delta R.T. 0.01 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Resp	Lower	Upper
75	100	437757		
110	37.4	18.1	54.1	
77	31.4	24.5	36.7	









#37
Ethyl Acetate
Concen: 51.788 ug/l
RT: 4.82 min Scan# 645
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampled :
VSTDCCC050EC

Tot Ion:	43	Resp:	528403
Ion	Ratio	Lower	Upper
43	100		
61	13.5	11.0	16.4
70	11.0	8.6	12.8




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#38
Carbon Tetrachloride
Concen: 49.917 ug/l
RT: 5.78 min Scan# 802
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion:	117	Resp:	418630
Ion	Ratio	Lower	Upper
117	100		
119	94.8	77.5	116.3
121	30.2	25.0	37.6









#39
Methvlcvclohexane
Concen: 47.928 ug/l
RT: 7.46 min Scan# 1078
Delta R.T. 0.01 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
83	100		
55	75.9	61.9	92.9
98	48.2	39.4	59.0




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#40
Benzene
Concen: 49.139 ug/l
RT: 6.14 min Scan# 861
Delta R.T. 0.01 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
78	100		
77	24.0	19.0	28.4









#41
Methacrylonitrile
Concen: 52.233 ug/l
RT: 5.03 min Scan# 680
Delta R.T. 0.01 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ion	Ratio	Lower	Upper
292281	41	100		
	39	54.0	42.9	64.3
	67	72.0	57.4	86.0
	52	28.4	22.4	33.6




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#42
1,2-Dichloroethane
Concen: 49.902 ug/l
RT: 6.19 min Scan# 869
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ion	Ratio	Lower	Upper
471424	62	100		
	98	10.2	0.0	20.8









#43
Isopropyl Acetate
Concen: 50.140 ug/l
RT: 6.43 min Scan# 910
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
43	100		
61	20.4	16.3	24.5
87	13.4	10.8	16.2




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#44
Trichloroethene
Concen: 48.036 ug/l
RT: 7.21 min Scan# 1037
Delta R.T. 0.01 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
130	100		
95	88.4	0.0	190.8









#45
 1,2-Dichloropropane
 Concen: 51.106 ug/l
 RT: 7.51 min Scan# 1086
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
63	100		
65	31.0	25.5	38.3




Manual Integrations
 APPROVED

apatel
 12/2/2019 3:18:17 PM



#46
 Dibromomethane
 Concen: 48.887 ug/l
 RT: 7.65 min Scan# 1110
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
93	100		
95	83.1	66.2	99.2
174	116.6	93.4	140.0









#47
Bromodichloromethane
Concen: 50.141 ug/l
RT: 7.89 min Scan# 1149
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
83	100		
85	64.9	52.1	78.1
127	8.7	7.0	10.4




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#48
Methyl methacrylate
Concen: 52.068 ug/l
RT: 7.76 min Scan# 1128
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
41	100		
69	83.2	64.7	97.1
39	55.1	43.8	65.6









#49
1,4-Dioxane
Concen: 1007.073 ug/l
RT: 7.73 min Scan# 1123
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
88	100		
43	32.2	25.8	38.6
58	71.6	54.6	82.0




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#50
Toluene-d8
Concen: 48.584 ug/l
RT: 8.71 min Scan# 1283
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
98	100		
100	66.2	53.1	79.7





#51

4-Methyl-2-Pentanone

Concen: 259.575 ug/l

RT: 8.64 min Scan# 1271

Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tot Ion: 43 Resp: 2742736

Ion Ratio Lower Upper

43 100





58 39.9 32.1 48.1

Manual Integrations

APPROVED

apatel

12/2/2019 3:18:17 PM



#52

Toluene

Concen: 49.389 ug/l

RT: 8.78 min Scan# 1295

Delta R.T. 0.00 min

Lab File: VX013665.D




Acq: 27 Nov 2019 20:31


Tgt Ion: 92 Resp: 849919

Ion Ratio Lower Upper

92 100

91 172.3 136.2 204.2









#53
t-1,3-Dichloropropene
Concen: 49.106 ug/l
RT: 9.04 min Scan# 1337
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
75	100		
77	32.6	24.4	36.6




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#54
cis-1,3-Dichloropropene
Concen: 50.210 ug/l
RT: 8.43 min Scan# 1237
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
75	100		
77	31.7	25.8	38.6
39	52.1	39.9	59.9









#55
 1,1,2-Trichloroethane
 Concen: 50.615 ug/l
 RT: 9.21 min Scan# 1365
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion:	97	Resp:	351005
Ion	Ratio	Lower	Upper
97	100		
83	83.3	67.3	100.9
85	55.1	43.6	65.4
99	61.7	48.2	72.4




Manual Integrations
 APPROVED


apatel
 12/2/2019 3:18:17 PM



#56
 Ethyl methacrylate
 Concen: 51.462 ug/l
 RT: 9.17 min Scan# 1359
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Tgt Ion:	69	Resp:	573766
Ion	Ratio	Lower	Upper
69	100		
41	67.8	55.0	82.4
39	35.8	28.7	43.1









#57
 1,3-Dichloropropane
 Concen: 49.201 ug/l
 RT: 9.37 min Scan# 1391
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
76	100		
78	32.3	26.0	39.0




Manual Integrations
 APPROVED


apatel
 12/2/2019 3:18:17 PM



#58
 2-Chloroethyl Vinyl ether
 Concen: 255.251 ug/l
 RT: 8.31 min Scan# 1217
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
63	100		
106	29.3	23.4	35.0









#59
2-Hexanone
Concen: 254.816 ug/l
RT: 9.48 min Scan# 1410
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion: 43 Resp: 2099907
Ion Ratio Lower Upper
43 100
58 55.4 28.1 84.4




Manual Integrations
APPROVED

apatel
12/2/2019 3:18:17 PM



#60
Dibromochloromethane
Concen: 50.458 ug/l
RT: 9.58 min Scan# 1426
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion: 129 Resp: 363143
Ion Ratio Lower Upper
129 100
127 77.7 38.5 115.5









#61
 1,2-Dibromoethane
 Concen: 49.293 ug/l
 RT: 9.67 min Scan# 1440
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
107	100		
109	93.3	76.1	114.1




Manual Integrations
 APPROVED


apatel
 12/2/2019 3:18:17 PM



#62
 4-Bromofluorobenzene
 Concen: 48.234 ug/l
 RT: 11.14 min Scan# 1681
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
95	100		
174	88.3	0.0	180.0
176	86.8	0.0	173.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Resp Lower	Upper
117	100		
82	52.8	41.8	62.8
119	32.9	25.8	38.8




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#64
Tetrachloroethene
Concen: 47.408 ug/l
RT: 9.33 min Scan# 1385
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Resp Lower	Upper
164	100		
166	129.8	99.7	149.5
129	90.9	70.1	105.1
131	85.1	66.9	100.3









#65
Chlorobenzene
Concen: 48.491 ug/l
RT: 10.14 min Scan# 1517
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
112	100		
114	32.4	26.1	39.1




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#66
1,1,1,2-Tetrachloroethane
Concen: 50.339 ug/l
RT: 10.21 min Scan# 1530
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
131	100		
133	95.7	47.7	143.1
119	65.7	32.9	98.6









#67
Ethyl Benzene
Concen: 50.322 ug/l
RT: 10.25 min Scan# 1535
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion: 91 Resp: 1626246
Ion Ratio Lower Upper
91 100
106 31.1 24.5 36.7




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#68
m/p-Xylenes
Concen: 99.768 ug/l
RT: 10.36 min Scan# 1553
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion: 106 Resp: 1234393
Ion Ratio Lower Upper
106 100
91 199.7 159.0 238.4









#69
o-Xylene
Concen: 49.766 ug/l
RT: 10.70 min Scan# 1609
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
106	100		
91	208.9	104.1	312.2




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#70
Styrene
Concen: 50.567 ug/l
RT: 10.71 min Scan# 1611
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
104	100		
78	48.4	38.5	57.7
103	54.6	43.7	65.5









#71
Bromoform
Concen: 49.976 ug/l
RT: 10.85 min Scan# 1635
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
173	100		
175	49.2	24.3	72.8
254	0.3	0.2	0.4




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
152	100		
115	77.7	38.4	115.1
150	174.8	0.0	346.8





#73

Isopropylbenzene

Concen: 49.403 ug/l

RT: 11.01 min Scan# 1661

Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tot Ion: 105 Resp: 1572029

Ion Ratio Lower Upper

105 100





120 27.5 13.5 40.5

Manual Integrations

APPROVED

apatel

12/2/2019 3:18:17 PM



#74

N-ethyl acetate

Concen: 51.476 ug/l

RT: 10.89 min Scan# 1641

Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Tgt Ion: 43 Resp: 750344




Ion Ratio Lower Upper


43 100

70 43.4 34.3 51.5

55 23.7 18.5 27.7

61 24.8 19.8 29.8





#75

1,1,2,2-Tetrachloroethane

Concen: 50.692 ug/l

RT: 11.26 min Scan# 1702

Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050EC

Tot Ion:	83	Resp:	559401
Ion Ratio	Lower	Upper	
83	100		
131	9.8	5.1	15.3
85	64.2	31.8	95.3

Manual Integrations
APPROVED

apatel
12/2/2019 3:18:17 PM



#76

1,2,3-Trichloropropane

Concen: 43.090 ug/l m




RT: 11.29 min Scan# 1707


Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Tgt Ion:	75	Resp:	439652
Ion Ratio	Lower	Upper	
75	100		
77	44.3	19.1	57.5





#77

Bromobenzene

Concen: 48.203 ug/l

RT: 11.25 min Scan# 1700

Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Instrument :

MSVOA_X





ClientSampleId :

VSTDCCC050EC

Tot Ion:	156	Resp:	414373
Ion Ratio	Lower	Upper	
156	100		
77	151.2	74.4	223.1
158	98.8	48.9	146.6

Manual Integrations
APPROVED

apatel
12/2/2019 3:18:17 PM



#78

n-propylbenzene

Concen: 49.860 ug/l




RT: 11.35 min Scan# 1717


Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Tgt Ion:	91	Resp:	1771102
Ion Ratio	Lower	Upper	
91	100		
120	23.8	11.8	35.3









#79
 2-Chlorotoluene
 Concen: 49.428 ug/l
 RT: 11.42 min Scan# 1727
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion: 91 Resp: 1053226
 Ion Ratio Lower Upper
 91 100
 126 35.1 17.5 52.5




Manual Integrations
 APPROVED


apatel
 12/2/2019 3:18:17 PM



#80
 1,3,5-Trimethylbenzene
 Concen: 50.636 ug/l
 RT: 11.50 min Scan# 1741
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Tgt Ion: 105 Resp: 1320579
 Ion Ratio Lower Upper
 105 100
 120 50.5 25.3 75.9









#81
trans-1,4-Dichloro-2-butene
Concen: 48.473 ug/l
RT: 11.07 min Scan# 1670
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
75	100		
53	96.4	78.2	117.2
89	45.2	35.4	53.2




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#82
4-Chlorotoluene
Concen: 49.245 ug/l
RT: 11.51 min Scan# 1742
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
91	100		
126	31.0	15.7	47.0









#83
tert-Butylbenzene
Concen: 50.415 ug/l
RT: 11.76 min Scan# 1784
Delta R.T. -0.01 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
119	100		
91	53.6	27.2	81.6
134	23.5	11.7	35.1




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#84
1,2,4-Trimethylbenzene
Concen: 50.195 ug/l
RT: 11.81 min Scan# 1791
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
105	100		
120	47.1	23.1	69.3





#85

sec-Butylbenzene

Concen: 50.366 ug/l

RT: 11.94 min Scan# 1813

Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Instrument :

MSVOA_X

ClientSampleID :

VSTDCCC050EC

Tot Ion:105 Resp: 1524781

Ion Ratio Lower Upper


105 100

134 20.6 10.4 31.2

Manual Integrations
APPROVED




apatel

12/2/2019 3:18:17 PM



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 50.087 ug/l

RT: 12.06 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31


Tgt Ion:119 Resp: 1410659

Ion Ratio Lower Upper

119 100

134 27.0 13.3 39.9



91 22.2 11.3 33.8




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): V









#87
 1,3-Dichlorobenzene
 Concen: 47.792 ug/l
 RT: 12.02 min Scan# 1826
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
146	100		
111	38.0	19.0	57.0
148	64.6	31.8	95.4




Manual Integrations
 APPROVED


apatel
 12/2/2019 3:18:17 PM



#88
 1,4-Dichlorobenzene
 Concen: 46.654 ug/l
 RT: 12.09 min Scan# 1838
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
146	100		
111	37.6	18.9	56.9
148	64.5	31.9	95.5









#89
n-Butylbenzene
Concen: 50.034 ug/l
RT: 12.39 min Scan# 1886
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
91	100		
92	54.7	27.4	82.2
134	27.3	13.5	40.4




Manual Integrations
APPROVED


apatel
12/2/2019 3:18:17 PM



#90
Hexachloroethane
Concen: 49.272 ug/l
RT: 12.59 min Scan# 1920
Delta R.T. 0.00 min
Lab File: VX013665.D
Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
117	100		
201	87.1	44.1	132.4









#91
 1,2-Dichlorobenzene
 Concen: 48.732 ug/l
 RT: 12.39 min Scan# 1886
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
146	100		
111	39.8	19.6	58.7
148	64.4	31.9	95.7




Manual Integrations
 APPROVED


apatel
 12/2/2019 3:18:17 PM



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 50.981 ug/l
 RT: 12.99 min Scan# 1986
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Lower	Upper
75	100		
155	97.1	47.6	142.8
157	125.6	62.1	186.2









#93
 1,2,4-Trichlorobenzene
 Concen: 47.226 ug/l
 RT: 13.64 min Scan# 2092
 Delta R.T. 0.00 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion	Ratio	Resp	Lower	Upper
180	100	487304		
182	97.0	47.2	141.6	
145	28.4	13.8	41.4	




Manual Integrations
 APPROVED


apatel
 12/2/2019 3:18:17 PM



#94
 Hexachlorobutadiene
 Concen: 45.532 ug/l
 RT: 13.78 min Scan# 2115
 Delta R.T. 0.01 min
 Lab File: VX013665.D
 Acq: 27 Nov 2019 20:31

Tgt Ion	Ratio	Resp	Lower	Upper
225	100	225693		
223	64.3	31.9	95.5	
227	64.2	32.1	96.3	





#95

Naphthalene

Concen: 52.060 ug/l

RT: 13.83 min Scan# 2123

Delta R.T. 0.00 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tot Ion:128 Resp: 1599758

Ion Ratio Lower Upper

128 100





127 13.0 10.3 15.5

129 11.0 8.7 13.1

Manual Integrations
APPROVED

apatel

12/2/2019 3:18:17 PM



#96

1,2,3-Trichlorobenzene

Concen: 49.018 ug/l

RT: 14.01 min Scan# 2153

Delta R.T. -0.01 min

Lab File: VX013665.D

Acq: 27 Nov 2019 20:31

Tgt Ion:180 Resp: 500800

Ion Ratio Lower Upper

180 100

182 97.2 47.6 142.8

145 29.9 15.0 45.0

