

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112723\
 Data File : VX039033.D
 Acq On : 27 Nov 2023 11:41
 Operator : JC/MD
 Sample : VX1127WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1127WBL01

Quant Time: Nov 28 01:11:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Nov 27 03:22:25 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John Carlone 11/28/2023
 Supervised By :Mahesh Dadoda 11/28/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.910	128	9668m	30.000	ug/l	0.01
28) 1,4-Difluorobenzene	6.763	114	79743	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	96200	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	43467	32.864	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	= 109.533%	
60) 4-Bromofluorobenzene	11.079	95	60235	28.667	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	= 95.567%	
63) Toluene-d8	8.647	98	116250	30.565	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	= 101.867%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112723\
Data File : VX039033.D
Acq On : 27 Nov 2023 11:41
Operator : JC/MD
Sample : VX1127WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1127WBL01

Manual Integrations
APPROVED

Reviewed By : John Carlone 11/28/2023
Supervised By : Mahesh Dadoda 11/28/2023

Quant Time: Nov 28 01:11:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Nov 27 03:22:25 2023
Response via : Initial Calibration

