

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112723\
 Data File : VX039037.D
 Acq On : 27 Nov 2023 13:32
 Operator : JC/MD
 Sample : 05525-01RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TWP10RE

Quant Time: Nov 28 01:12:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Nov 27 03:22:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	9261	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	75930	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	92266	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	42135	33.257	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	110.867%#
60) 4-Bromofluorobenzene	11.079	95	57226	28.396	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	94.667%
63) Toluene-d8	8.647	98	114005	31.252	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	104.167%
Target Compounds						
					Qvalue	
15) Acetone	2.379	58	492	1.409	ug/l #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112723\
Data File : VX039037.D
Acq On : 27 Nov 2023 13:32
Operator : JC/MD
Sample : 05525-01RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TWP10RE

Quant Time: Nov 28 01:12:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Nov 27 03:22:25 2023
Response via : Initial Calibration

