

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112723\
 Data File : VX039041.D
 Acq On : 27 Nov 2023 15:23
 Operator : JC/MD
 Sample : 05525-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TWP14

Quant Time: Nov 28 01:13:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Nov 27 03:22:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	9355	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	70898	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	88526	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	40559	31.691	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	105.633%
60) 4-Bromofluorobenzene	11.079	95	56170	29.050	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	96.833%
63) Toluene-d8	8.647	98	106682	30.480	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.600%
Target Compounds						
15) Acetone	2.380	58	803	2.276	ug/l #	Qvalue 50

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112723\
Data File : VX039041.D
Acq On : 27 Nov 2023 15:23
Operator : JC/MD
Sample : 05525-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TWP14

Quant Time: Nov 28 01:13:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Nov 27 03:22:25 2023
Response via : Initial Calibration

