

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
 Data File : VX044025.D
 Acq On : 27 Nov 2024 09:15
 Operator : JC/MD
 Sample : VX1127MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1127MBL01

Quant Time: Nov 28 00:39:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	119396	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	228047	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	207542	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	87354	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	100927	49.284	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.560%
35) Dibromofluoromethane	5.379	113	76981	47.242	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.480%
50) Toluene-d8	8.647	98	280743	49.980	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.960%
62) 4-Bromofluorobenzene	11.079	95	94181	49.267	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.540%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
Data File : VX044025.D
Acq On : 27 Nov 2024 09:15
Operator : JC/MD
Sample : VX1127MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1127MBL01

Quant Time: Nov 28 00:39:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

