

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
 Data File : VX044029.D
 Acq On : 27 Nov 2024 10:50
 Operator : JC/MD
 Sample : P5006-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-WTS-02

Quant Time: Nov 28 00:42:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	110847	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	211591	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	195539	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	80518	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	97214	51.133	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.260%
35) Dibromofluoromethane	5.379	113	71633	47.379	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.760%
50) Toluene-d8	8.646	98	270234	51.851	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.700%
62) 4-Bromofluorobenzene	11.079	95	86685	48.872	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.740%
Target Compounds						
					Qvalue	
16) Acetone	2.391	43	1399	1.602	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
Data File : VX044029.D
Acq On : 27 Nov 2024 10:50
Operator : JC/MD
Sample : P5006-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-WTS-02

Quant Time: Nov 28 00:42:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

