

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
 Data File : VX044030.D
 Acq On : 27 Nov 2024 11:13
 Operator : JC/MD
 Sample : PB165270TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB165270TB

Quant Time: Nov 28 00:42:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	112058	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	215861	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191823	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	82777	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	99294	51.662	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.320%
35) Dibromofluoromethane	5.379	113	71894	46.611	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.220%
50) Toluene-d8	8.647	98	264151	49.681	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.360%
62) 4-Bromofluorobenzene	11.079	95	90089	49.787	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.580%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
Data File : VX044030.D
Acq On : 27 Nov 2024 11:13
Operator : JC/MD
Sample : PB165270TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB165270TB

Quant Time: Nov 28 00:42:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

