

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
 Data File : VX044032.D
 Acq On : 27 Nov 2024 11:59
 Operator : JC/MD
 Sample : P5001-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C1-2

Quant Time: Nov 29 03:32:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	123309	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	238082	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208221	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98083	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	106185	50.207	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	100.420%	
35) Dibromofluoromethane	5.385	113	57505	33.802	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	67.600%#	
50) Toluene-d8	8.646	98	289748	49.409	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.820%	
62) 4-Bromofluorobenzene	11.079	95	100482	50.347	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	100.700%	
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	9563	9.845	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	4707	2.596	ug/l	87
40) Benzene	6.031	78	155986	23.622	ug/l	99
52) Toluene	8.713	92	223683	56.595	ug/l	99
67) Ethyl Benzene	10.189	91	1057886	136.516	ug/l	99
68) m/p-Xylenes	10.299	106	191715	66.341	ug/l	100
69) o-Xylene	10.640	106	130499	46.251	ug/l	100
73) Isopropylbenzene	10.963	105	45933	6.092	ug/l	100
78) n-propylbenzene	11.304	91	10304	1.224	ug/l #	53
80) 1,3,5-Trimethylbenzene	11.451	105	41534	6.696	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	154496	25.018	ug/l	99
95) Naphthalene	13.780	128	12258842	1747.332	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
Data File : VX044032.D
Acq On : 27 Nov 2024 11:59
Operator : JC/MD
Sample : P5001-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C1-2

Quant Time: Nov 29 03:32:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

