

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
 Data File : VX044033.D
 Acq On : 27 Nov 2024 12:22
 Operator : JC/MD
 Sample : P5001-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C1-3

Quant Time: Nov 28 00:44:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	121169	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	227299	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	203619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98952	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	102751	49.441	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.880%
35) Dibromofluoromethane	5.385	113	51191	31.518	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	63.040%#
50) Toluene-d8	8.647	98	280952	50.182	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.360%
62) 4-Bromofluorobenzene	11.079	95	99073	51.996	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.000%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	9871	10.341	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	1948	1.093	ug/l	97
40) Benzene	6.031	78	64688	10.261	ug/l	98
52) Toluene	8.714	92	157506	41.742	ug/l	100
67) Ethyl Benzene	10.189	91	964297	127.251	ug/l	98
68) m/p-Xylenes	10.299	106	176821	62.570	ug/l	97
69) o-Xylene	10.640	106	120204	43.565	ug/l	96
73) Isopropylbenzene	10.963	105	46801	6.153	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	44141	7.054	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	157988	25.358	ug/l	100
95) Naphthalene	13.780	128	12832701	1813.064	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
Data File : VX044033.D
Acq On : 27 Nov 2024 12:22
Operator : JC/MD
Sample : P5001-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C1-3

Quant Time: Nov 28 00:44:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

