

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
 Data File : VX044036.D
 Acq On : 27 Nov 2024 13:31
 Operator : JC/MD
 Sample : P5001-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C2-3

Quant Time: Nov 29 04:09:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	117897	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	224840	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	200377	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	94742	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	100646	49.772	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	99.540%	
35) Dibromofluoromethane	5.385	113	49501	30.811	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	61.620%#	
50) Toluene-d8	8.647	98	279518	50.472	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	100.940%	
62) 4-Bromofluorobenzene	11.079	95	98104	52.051	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	104.100%	
Target Compounds						Qvalue
4) Vinyl Chloride	1.374	62	9957	5.956	ug/l	94
16) Acetone	2.386	43	17142	18.457	ug/l #	89
27) cis-1,2-Dichloroethene	4.483	96	270698	156.136	ug/l	91
40) Benzene	6.031	78	640238	102.666	ug/l	98
44) Trichloroethene	7.116	130	29866	16.918	ug/l	94
52) Toluene	8.714	92	533116	142.830	ug/l	99
67) Ethyl Benzene	10.189	91	3226170	432.623	ug/l	98
68) m/p-Xylenes	10.299	106	466477	167.739	ug/l	99
69) o-Xylene	10.640	106	366257	134.890	ug/l	99
73) Isopropylbenzene	10.963	105	127113	17.454	ug/l	98
78) n-propylbenzene	11.305	91	31747	3.903	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	114264	19.071	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	424310	71.131	ug/l	99
86) p-Isopropyltoluene	12.006	119	17535	2.967	ug/l #	74
95) Naphthalene	13.780	128	18284864	2698.167	ug/l #	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
Data File : VX044036.D
Acq On : 27 Nov 2024 13:31
Operator : JC/MD
Sample : P5001-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C2-3

Quant Time: Nov 29 04:09:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

