

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
 Data File : VX044037.D
 Acq On : 27 Nov 2024 13:54
 Operator : JC/MD
 Sample : P5001-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C3-1

Quant Time: Nov 28 00:46:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	117909	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	226519	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	201052	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	94799	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	100307	49.599	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	99.200%	
35) Dibromofluoromethane	5.385	113	48337	29.864	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	59.720%#	
50) Toluene-d8	8.647	98	275771	49.426	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.860%	
62) 4-Bromofluorobenzene	11.079	95	100072	52.702	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	105.400%	
Target Compounds					Qvalue	
16) Acetone	2.380	43	16137	17.373	ug/l	98
27) cis-1,2-Dichloroethene	4.483	96	13063	7.534	ug/l	88
40) Benzene	6.031	78	17146	2.729	ug/l	98
44) Trichloroethene	7.129	130	3809	2.142	ug/l	73
52) Toluene	8.714	92	32752	8.710	ug/l	99
67) Ethyl Benzene	10.189	91	252563	33.754	ug/l	100
68) m/p-Xylenes	10.299	106	43510	15.593	ug/l	99
69) o-Xylene	10.640	106	36898	13.544	ug/l	96
73) Isopropylbenzene	10.963	105	16225	2.227	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	16817	2.805	ug/l	97
84) 1,2,4-Trimethylbenzene	11.750	105	62598	10.488	ug/l	100
95) Naphthalene	13.774	128	6055671	893.055	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
Data File : VX044037.D
Acq On : 27 Nov 2024 13:54
Operator : JC/MD
Sample : P5001-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C3-1

Quant Time: Nov 28 00:46:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

