

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
 Data File : VX044038.D
 Acq On : 27 Nov 2024 14:16
 Operator : JC/MD
 Sample : P5001-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C3-2

Quant Time: Nov 28 00:47:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	118131	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	228473	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	200936	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	95305	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	101796	50.241	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	100.480%	
35) Dibromofluoromethane	5.379	113	48468	29.688	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	59.380%#	
50) Toluene-d8	8.647	98	279981	49.752	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	99.500%	
62) 4-Bromofluorobenzene	11.079	95	99175	51.783	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	103.560%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	43151	46.369	ug/l	97
25) 2-Butanone	4.574	43	8899	7.411	ug/l	98
27) cis-1,2-Dichloroethene	4.477	96	120308	69.255	ug/l	90
40) Benzene	6.031	78	586629	92.574	ug/l	99
44) Trichloroethene	7.123	130	3200	1.784	ug/l	81
52) Toluene	8.714	92	1041836	274.686	ug/l	99
67) Ethyl Benzene	10.195	91	3471652	464.246	ug/l	96
68) m/p-Xylenes	10.299	106	616308	221.000	ug/l	100
69) o-Xylene	10.640	106	411349	151.076	ug/l	96
73) Isopropylbenzene	10.957	105	130381	17.797	ug/l	99
78) n-propylbenzene	11.305	91	34425	4.207	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	126808	21.039	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	468891	78.141	ug/l	100
95) Naphthalene	13.780	128	20032822	2938.638	ug/l #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
Data File : VX044038.D
Acq On : 27 Nov 2024 14:16
Operator : JC/MD
Sample : P5001-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C3-2

Quant Time: Nov 28 00:47:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

