

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
 Data File : VX044039.D
 Acq On : 27 Nov 2024 14:39
 Operator : JC/MD
 Sample : P5001-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C3-3

Quant Time: Nov 28 00:48:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	T.Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) Pentafluorobenzene	5.549	168		114924	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114		221134	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117		195996	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152		90506	50.000	ug/l	0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4	5.952	65		101369	51.427	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125		Recovery	=	102.860%	
35) Dibromofluoromethane	5.379	113		47618	30.136	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124		Recovery	=	60.280%#	
50) Toluene-d8	8.646	98		274097	50.322	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113		Recovery	=	100.640%	
62) 4-Bromofluorobenzene	11.079	95		96086	51.835	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121		Recovery	=	103.660%	
Target Compounds							Qvalue
16) Acetone	2.385	43		10943	12.087	ug/l	89
27) cis-1,2-Dichloroethene	4.483	96		13439	7.952	ug/l	91
40) Benzene	6.031	78		319156	52.036	ug/l	98
52) Toluene	8.713	92		272927	74.347	ug/l	99
67) Ethyl Benzene	10.189	91		2313152	317.122	ug/l	99
68) m/p-Xylenes	10.299	106		328414	120.733	ug/l	98
69) o-Xylene	10.640	106		265526	99.977	ug/l	97
73) Isopropylbenzene	10.957	105		98161	14.109	ug/l	98
78) n-propylbenzene	11.304	91		25637	3.299	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105		82555	14.423	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105		298619	52.404	ug/l	100
95) Naphthalene	13.780	128	15405454	2379.670	ug/l	#	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
Data File : VX044039.D
Acq On : 27 Nov 2024 14:39
Operator : JC/MD
Sample : P5001-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C3-3

Quant Time: Nov 28 00:48:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

