

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
 Data File : VX044040.D
 Acq On : 27 Nov 2024 15:02
 Operator : JC/MD
 Sample : P5001-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-C4-1

Quant Time: Nov 28 00:49:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	113725	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	222916	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	203256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	88732	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	101660	52.118	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.240%		
35) Dibromofluoromethane	5.379	113	52838	33.172	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 66.340%#		
50) Toluene-d8	8.647	98	277165	50.479	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.960%		
62) 4-Bromofluorobenzene	11.079	95	99567	53.283	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 106.560%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	15950	17.804	ug/l	95
27) cis-1,2-Dichloroethene	4.483	96	7515	4.494	ug/l	91
67) Ethyl Benzene	10.195	91	24818	3.281	ug/l	99
69) o-Xylene	10.640	106	3325	1.207	ug/l	95
84) 1,2,4-Trimethylbenzene	11.750	105	7482	1.339	ug/l	99
95) Naphthalene	13.774	128	824849	129.961	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
Data File : VX044040.D
Acq On : 27 Nov 2024 15:02
Operator : JC/MD
Sample : P5001-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-C4-1

Quant Time: Nov 28 00:49:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

