

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
 Data File : VX044024.D
 Acq On : 27 Nov 2024 08:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 28 00:38:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.633	0.577	8.8	79	0.00
3 P	Chloromethane	0.667	0.617	7.5	83	0.00
4 C	Vinyl Chloride	0.709	0.636	10.3#	79	0.00
5 T	Bromomethane	0.436	0.432	0.9	91	0.00
6 T	Chloroethane	0.432	0.428	0.9	88	0.00
7 T	Trichlorofluoromethane	1.268	1.126	11.2	78	0.00
8 T	Diethyl Ether	0.462	0.414	10.4	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.605	0.569	6.0	81	0.00
10 T	Methyl Iodide	0.755	0.711	5.8	83	0.00
11 T	Tert butyl alcohol	0.138	0.125	9.4	81	0.00
12 CM	1,1-Dichloroethene	0.576	0.512	11.1#	80	0.00
13 T	Acrolein	0.145	0.118	18.6	77	0.00
14 T	Allyl chloride	0.941	0.878	6.7	79	0.00
15 T	Acrylonitrile	0.335	0.319	4.8	81	0.00
16 T	Acetone	0.394	0.390	1.0	86	0.00
17 T	Carbon Disulfide	1.188	0.969	18.4	71	0.00
18 T	Methyl Acetate	0.917	0.850	7.3	81	0.00
19 T	Methyl tert-butyl Ether	2.092	1.967	6.0	82	0.00
20 T	Methylene Chloride	0.668	0.602	9.9	85	0.00
21 T	trans-1,2-Dichloroethene	0.595	0.533	10.4	79	0.00
22 T	Diisopropyl ether	2.013	1.854	7.9	81	0.00
23 T	Vinyl Acetate	1.722	1.649	4.2	80	0.00
24 P	1,1-Dichloroethane	1.161	1.078	7.1	81	0.00
25 T	2-Butanone	0.508	0.489	3.7	81	0.00
26 T	2,2-Dichloropropane	1.049	0.976	7.0	81	0.00
27 T	cis-1,2-Dichloroethene	0.735	0.673	8.4	80	0.00
28 T	Bromochloromethane	0.511	0.491	3.9	103	0.00
29 T	Tetrahydrofuran	0.312	0.286	8.3	80	0.00
30 C	Chloroform	1.249	1.167	6.6#	83	0.00
31 T	Cyclohexane	0.956	0.850	11.1	77	0.00
32 T	1,1,1-Trichloroethane	1.077	1.030	4.4	82	0.00
33 S	1,2-Dichloroethane-d4	0.858	0.810	5.6	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.357	0.370	-3.6	103	0.00
36 T	1,1-Dichloropropene	0.445	0.428	3.8	79	0.00
37 T	Ethyl Acetate	0.553	0.523	5.4	78	0.00
38 T	Carbon Tetrachloride	0.500	0.512	-2.4	83	0.00
39 T	Methylcyclohexane	0.578	0.559	3.3	77	0.00
40 TM	Benzene	1.387	1.345	3.0	82	0.00
41 T	Methacrylonitrile	0.283	0.286	-1.1	79	0.00
42 TM	1,2-Dichloroethane	0.557	0.536	3.8	81	0.00
43 T	Isopropyl Acetate	0.880	0.861	2.2	80	0.00
44 TM	Trichloroethene	0.393	0.337	14.2	82	0.00
45 C	1,2-Dichloropropane	0.340	0.345	-1.5#	84	0.00
46 T	Dibromomethane	0.269	0.274	-1.9	85	0.00
47 T	Bromodichloromethane	0.482	0.503	-4.4	84	0.00
48 T	Methyl methacrylate	0.422	0.410	2.8	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
 Data File : VX044024.D
 Acq On : 27 Nov 2024 08:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 28 00:38:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	80	0.00
50 S	Toluene-d8	1.232	1.254	-1.8	103	0.00
51 T	4-Methyl-2-Pentanone	0.537	0.532	0.9	79	0.00
52 CM	Toluene	0.830	0.830	0.0	82	0.00
53 T	t-1,3-Dichloropropene	0.491	0.502	-2.2	80	0.00
54 T	cis-1,3-Dichloropropene	0.539	0.543	-0.7	81	0.00
55 T	1,1,2-Trichloroethane	0.343	0.335	2.3	82	0.00
56 T	Ethyl methacrylate	0.534	0.530	0.7	78	0.00
57 T	1,3-Dichloropropane	0.577	0.579	-0.3	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.273	0.274	-0.4	81	0.00
59 T	2-Hexanone	0.403	0.411	-2.0	79	0.00
60 T	Dibromochloromethane	0.344	0.361	-4.9	83	0.00
61 T	1,2-Dibromoethane	0.346	0.349	-0.9	82	0.00
62 S	4-Bromofluorobenzene	0.419	0.419	0.0	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.330	0.338	-2.4	84	0.00
65 PM	Chlorobenzene	1.098	1.077	1.9	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.356	0.377	-5.9	83	0.00
67 C	Ethyl Benzene	1.861	1.877	-0.9#	81	0.00
68 T	m/p-Xylenes	0.694	0.703	-1.3	80	0.00
69 T	o-Xylene	0.678	0.683	-0.7	79	0.00
70 T	Styrene	1.121	1.141	-1.8	79	0.00
71 P	Bromoform	0.240	0.257	-7.1	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
73 T	Isopropylbenzene	3.843	3.898	-1.4	82	0.00
74 T	N-amyl acetate	1.780	1.742	2.1	75	0.00
75 P	1,1,2,2-Tetrachloroethane	1.316	1.301	1.1	80	0.00
76 T	1,2,3-Trichloropropane	1.096	1.081	1.4	79	0.00
77 T	Bromobenzene	0.917	0.912	0.5	81	0.00
78 T	n-propylbenzene	4.293	4.376	-1.9	80	0.00
79 T	2-Chlorotoluene	2.714	2.638	2.8	80	0.00
80 T	1,3,5-Trimethylbenzene	3.162	3.201	-1.2	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.402	0.385	4.2	75	0.00
82 T	4-Chlorotoluene	3.064	3.052	0.4	79	0.00
83 T	tert-Butylbenzene	3.134	3.187	-1.7	80	0.00
84 T	1,2,4-Trimethylbenzene	3.148	3.189	-1.3	80	0.00
85 T	sec-Butylbenzene	3.840	3.885	-1.2	79	0.00
86 T	p-Isopropyltoluene	3.119	3.286	-5.4	80	0.00
87 T	1,3-Dichlorobenzene	1.670	1.662	0.5	81	0.00
88 T	1,4-Dichlorobenzene	1.702	1.645	3.3	80	0.00
89 T	n-Butylbenzene	2.750	2.884	-4.9	78	0.00
90 T	Hexachloroethane	0.530	0.533	-0.6	78	0.00
91 T	1,2-Dichlorobenzene	1.685	1.674	0.7	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.264	0.263	0.4	80	0.00
93 T	1,2,4-Trichlorobenzene	0.994	0.984	1.0	78	0.00
94 T	Hexachlorobutadiene	0.393	0.389	1.0	81	0.00
95 T	Naphthalene	3.576	3.607	-0.9	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
Data File : VX044024.D
Acq On : 27 Nov 2024 08:46
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 28 00:38:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.006	1.025	-1.9	79	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5