

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
 Data File : VX044050.D
 Acq On : 27 Nov 2024 18:52
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 28 00:56:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	50.000	47.528	4.9	73	0.00
3 P	Chloromethane	50.000	45.475	9.0	72	0.00
4 C	Vinyl Chloride	50.000	46.471	7.1#	71	0.00
5 T	Bromomethane	50.000	53.949	-7.9	87	0.00
6 T	Chloroethane	50.000	50.675	-1.3	79	0.00
7 T	Trichlorofluoromethane	50.000	57.957	-15.9	89	0.00
8 T	Diethyl Ether	50.000	51.363	-2.7	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	52.377	-4.8	79	0.00
10 T	Methyl Iodide	50.000	51.356	-2.7	79	0.00
11 T	Tert butyl alcohol	250.000	238.988	4.4	75	0.00
12 CM	1,1-Dichloroethene	50.000	51.164	-2.3#	81	0.00
13 T	Acrolein	250.000	312.051	-24.8	103	0.00
14 T	Allyl chloride	50.000	53.612	-7.2	80	0.00
15 T	Acrylonitrile	250.000	259.344	-3.7	78	0.00
16 T	Acetone	250.000	217.568	13.0	67	0.00
17 T	Carbon Disulfide	50.000	42.897	14.2	65	0.00
18 T	Methyl Acetate	50.000	53.320	-6.6	82	0.00
19 T	Methyl tert-butyl Ether	50.000	57.645	-15.3	88	0.00
20 T	Methylene Chloride	50.000	50.191	-0.4	83	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.687	-3.4	80	0.00
22 T	Diisopropyl ether	50.000	56.815	-13.6	87	0.00
23 T	Vinyl Acetate	250.000	282.049	-12.8	83	0.00
24 P	1,1-Dichloroethane	50.000	55.287	-10.6	84	0.00
25 T	2-Butanone	250.000	250.578	-0.2	74	0.00
26 T	2,2-Dichloropropane	50.000	51.948	-3.9	79	0.00
27 T	cis-1,2-Dichloroethene	50.000	54.243	-8.5	83	0.00
28 T	Bromochloromethane	50.000	55.983	-12.0	105	0.00
29 T	Tetrahydrofuran	250.000	255.554	-2.2	78	0.00
30 C	Chloroform	50.000	56.152	-12.3#	88	0.00
31 T	Cyclohexane	50.000	50.238	-0.5	76	0.00
32 T	1,1,1-Trichloroethane	50.000	57.137	-14.3	86	0.00
33 S	1,2-Dichloroethane-d4	50.000	58.020	-16.0	109	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	87	0.00
35 S	Dibromofluoromethane	50.000	55.117	-10.2	105	0.00
36 T	1,1-Dichloropropene	50.000	52.169	-4.3	82	0.00
37 T	Ethyl Acetate	50.000	50.971	-1.9	81	0.00
38 T	Carbon Tetrachloride	50.000	52.261	-4.5	82	0.00
39 T	Methylcyclohexane	50.000	48.120	3.8	74	0.00
40 TM	Benzene	50.000	52.039	-4.1	84	0.00
41 T	Methacrylonitrile	50.000	53.731	-7.5	81	0.00
42 TM	1,2-Dichloroethane	50.000	54.863	-9.7	88	0.00
43 T	Isopropyl Acetate	50.000	54.423	-8.8	86	0.00
44 TM	Trichloroethene	50.000	43.859	12.3	81	0.00
45 C	1,2-Dichloropropane	50.000	54.002	-8.0#	86	0.00
46 T	Dibromomethane	50.000	53.829	-7.7	86	0.00
47 T	Bromodichloromethane	50.000	56.141	-12.3	87	0.00
48 T	Methyl methacrylate	50.000	54.301	-8.6	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
 Data File : VX044050.D
 Acq On : 27 Nov 2024 18:52
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 28 00:56:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1057.830	-5.8	85	0.00
50 S	Toluene-d8	50.000	54.924	-9.8	107	0.00
51 T	4-Methyl-2-Pentanone	250.000	269.752	-7.9	83	0.00
52 CM	Toluene	50.000	54.762	-9.5#	86	0.00
53 T	t-1,3-Dichloropropene	50.000	55.866	-11.7	84	0.00
54 T	cis-1,3-Dichloropropene	50.000	55.077	-10.2	85	0.00
55 T	1,1,2-Trichloroethane	50.000	53.779	-7.6	87	0.00
56 T	Ethyl methacrylate	50.000	55.750	-11.5	85	0.00
57 T	1,3-Dichloropropane	50.000	54.478	-9.0	88	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	250.399	-0.2	83	0.00
59 T	2-Hexanone	250.000	264.663	-5.9	79	0.00
60 T	Dibromochloromethane	50.000	55.278	-10.6	84	0.00
61 T	1,2-Dibromoethane	50.000	54.726	-9.5	86	0.00
62 S	4-Bromofluorobenzene	50.000	57.635	-15.3	109	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	85	0.00
64 T	Tetrachloroethene	50.000	50.521	-1.0	81	0.00
65 PM	Chlorobenzene	50.000	52.267	-4.5	85	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	55.875	-11.8	86	0.00
67 C	Ethyl Benzene	50.000	56.596	-13.2#	89	0.00
68 T	m/p-Xylenes	100.000	107.735	-7.7	83	0.00
69 T	o-Xylene	50.000	55.159	-10.3	84	0.00
70 T	Styrene	50.000	55.758	-11.5	84	0.00
71 P	Bromoform	50.000	47.597	4.8	77	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	87	0.00
73 T	Isopropylbenzene	50.000	52.833	-5.7	86	0.00
74 T	N-amyl acetate	50.000	56.617	-13.2	88	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	52.531	-5.1	86	0.00
76 T	1,2,3-Trichloropropane	50.000	50.505	-1.0	81	0.00
77 T	Bromobenzene	50.000	52.056	-4.1	85	0.00
78 T	n-propylbenzene	50.000	53.955	-7.9	85	0.00
79 T	2-Chlorotoluene	50.000	50.906	-1.8	84	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.321	-6.6	85	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.919	6.2	74	0.00
82 T	4-Chlorotoluene	50.000	53.478	-7.0	85	0.00
83 T	tert-Butylbenzene	50.000	54.725	-9.5	86	0.00
84 T	1,2,4-Trimethylbenzene	50.000	56.462	-12.9	90	0.00
85 T	sec-Butylbenzene	50.000	54.100	-8.2	85	0.00
86 T	p-Isopropyltoluene	50.000	54.835	-9.7	84	0.00
87 T	1,3-Dichlorobenzene	50.000	52.226	-4.5	86	0.00
88 T	1,4-Dichlorobenzene	50.000	50.664	-1.3	84	0.00
89 T	n-Butylbenzene	50.000	54.774	-9.5	82	0.00
90 T	Hexachloroethane	50.000	50.224	-0.4	78	0.00
91 T	1,2-Dichlorobenzene	50.000	52.516	-5.0	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.161	-0.3	81	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.779	0.4	79	0.00
94 T	Hexachlorobutadiene	50.000	48.084	3.8	80	0.00
95 T	Naphthalene	50.000	58.422	-16.8	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112724\
Data File : VX044050.D
Acq On : 27 Nov 2024 18:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 28 00:56:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.056	-4.1	81	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6