

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112822\
 Data File : VX033088.D
 Acq On : 28 Nov 2022 13:37
 Operator : JC/MD
 Sample : N5747-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-38D

Quant Time: Nov 29 01:43:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	131951	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	214832	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188515	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86385	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71477	55.799	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.600%
35) Dibromofluoromethane	5.391	113	56379	46.550	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.100%
50) Toluene-d8	8.647	98	191367	52.628	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.260%
62) 4-Bromofluorobenzene	11.079	95	69358	43.595	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.180%
Target Compounds						
					Qvalue	
67) Ethyl Benzene	10.195	91	2615	0.392	ug/l	94
68) m/p-Xylenes	10.311	106	1458	0.558	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	1523	0.294	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	4733	0.915	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112822\
Data File : VX033088.D
Acq On : 28 Nov 2022 13:37
Operator : JC/MD
Sample : N5747-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-38D

Quant Time: Nov 29 01:43:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

