

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112822\
 Data File : VX033098.D
 Acq On : 28 Nov 2022 18:33
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Nov 29 03:45:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X111022W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 11 05:24:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.909	128	10035	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	72083	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	75136	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	29467	30.438	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.467%
60) 4-Bromofluorobenzene	11.079	95	39062	27.178	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	90.600%
63) Toluene-d8	8.653	98	76019	29.980	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.933%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112822\
Data File : VX033098.D
Acq On : 28 Nov 2022 18:33
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Nov 29 03:45:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X111022W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 11 05:24:06 2022
Response via : Initial Calibration

