



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112822\
Data File : VX033079.D
Acq On : 28 Nov 2022 08:42
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Title : SW846 8260
Last Update : Tue Nov 22 13:58:46 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	8057	PASS
75	95	30	60	54.1	21931	PASS
95	95	100	100	100.0	40565	PASS
96	95	5	9	6.8	2760	PASS
173	174	0.00	2	1.2	382	PASS
174	95	50	100	80.5	32667	PASS
175	174	5	9	7.3	2399	PASS
176	174	95	101	95.6	31221	PASS
177	176	5	9	6.9	2158	PASS