



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112823\
Data File : VX039043.D
Acq On : 28 Nov 2023 09:07
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Mon Nov 27 03:22:25 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.6	3586	PASS
75	95	30	60	53.7	9352	PASS
95	95	100	100	100.0	17402	PASS
96	95	5	9	7.0	1221	PASS
173	174	0.00	2	0.5	63	PASS
174	95	50	100	68.0	11839	PASS
175	174	5	9	7.1	845	PASS
176	174	95	101	96.2	11390	PASS
177	176	5	9	6.6	750	PASS