

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX112921\  
 Data File : VX025366.D  
 Acq On : 29 Nov 2021 13:51  
 Operator : JC/MD  
 Sample : M4847-03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 STORAGE-BLANK-WATER-REF-4

Quant Time: Nov 30 04:01:24 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X111721W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Nov 17 15:36:16 2021  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.556	168	131747	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	211327	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191168	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86995	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70775	50.965	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.920%
35) Dibromofluoromethane	5.391	113	64356	49.633	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.260%
50) Toluene-d8	8.653	98	242296	49.700	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.400%
62) 4-Bromofluorobenzene	11.079	95	85564	46.393	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.780%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX112921\  
 Data File : VX025366.D  
 Acq On : 29 Nov 2021 13:51  
 Operator : JC/MD  
 Sample : M4847-03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 STORAGE-BLANK-WATER-REF-4

Quant Time: Nov 30 04:01:24 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X111721W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Nov 17 15:36:16 2021  
 Response via : Initial Calibration

