

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
 Data File : VX025370.D
 Acq On : 29 Nov 2021 15:24
 Operator : JC/MD
 Sample : M4816-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWBR3D-B-378-399-112221

Quant Time: Nov 29 15:55:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129566	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	208970	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182487	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78047	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69780	51.094	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.180%	
35) Dibromofluoromethane	5.391	113	64973	50.674	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.340%	
50) Toluene-d8	8.653	98	238906	49.557	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.120%	
62) 4-Bromofluorobenzene	11.079	95	79864	43.791	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 87.580%	
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.318	96	6882	5.731	ug/l	89
16) Acetone	2.379	43	30607	40.332	ug/l	94
21) trans-1,2-Dichloroethene	3.093	96	3939	2.994	ug/l	84
24) 1,1-Dichloroethane	3.611	63	5179	2.290	ug/l	96
25) 2-Butanone	4.568	43	6109	5.568	ug/l	93
27) cis-1,2-Dichloroethene	4.489	96	524478	341.393	ug/l	80
44) Trichloroethene	7.129	130	118810	74.052	ug/l	91
52) Toluene	8.720	92	27454	7.614	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
Data File : VX025370.D
Acq On : 29 Nov 2021 15:24
Operator : JC/MD
Sample : M4816-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWBR3D-B-378-399-112221

Quant Time: Nov 29 15:55:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

