

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
 Data File : VX025371.D
 Acq On : 29 Nov 2021 15:48
 Operator : JC/MD
 Sample : M4816-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-112221

Quant Time: Nov 29 16:15:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	130071	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206723	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186533	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84619	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69999	51.056	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.120%	
35) Dibromofluoromethane	5.391	113	63114	49.759	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.520%	
50) Toluene-d8	8.653	98	237440	49.789	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.580%	
62) 4-Bromofluorobenzene	11.085	95	83270	46.155	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.300%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
Data File : VX025371.D
Acq On : 29 Nov 2021 15:48
Operator : JC/MD
Sample : M4816-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-112221

Quant Time: Nov 29 16:15:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

