

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
 Data File : VX025372.D
 Acq On : 29 Nov 2021 16:11
 Operator : JC/MD
 Sample : M4816-03DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWBR3D-B-378-399-112221DL

Quant Time: Nov 29 16:35:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	208395	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187465	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84762	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70163	51.583	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.160%
35) Dibromofluoromethane	5.391	113	63355	49.549	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.100%
50) Toluene-d8	8.653	98	241041	50.138	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.280%
62) 4-Bromofluorobenzene	11.085	95	83792	46.072	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.140%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3514	4.649	ug/l	92
27) cis-1,2-Dichloroethene	4.489	96	54887	35.872	ug/l	80
44) Trichloroethene	7.129	130	12236	7.647	ug/l	93
52) Toluene	8.720	92	2869	0.798	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
Data File : VX025372.D
Acq On : 29 Nov 2021 16:11
Operator : JC/MD
Sample : M4816-03DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWBR3D-B-378-399-112221DL

Quant Time: Nov 29 16:35:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

