

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
 Data File : VX025373.D
 Acq On : 29 Nov 2021 16:34
 Operator : JC/MD
 Sample : M4816-02RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-112221RE

Quant Time: Nov 30 04:02:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127270	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206599	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83567	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70063	52.227	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.460%
35) Dibromofluoromethane	5.391	113	63200	49.857	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.720%
50) Toluene-d8	8.653	98	239781	50.310	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.620%
62) 4-Bromofluorobenzene	11.085	95	82942	46.001	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.000%
Target Compounds						
16) Acetone	2.380	43	1342	1.800	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
Data File : VX025373.D
Acq On : 29 Nov 2021 16:34
Operator : JC/MD
Sample : M4816-02RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-112221RE

Quant Time: Nov 30 04:02:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

