

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
 Data File : VX025374.D
 Acq On : 29 Nov 2021 16:58
 Operator : JC/MD
 Sample : M4837-11 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20931

Quant Time: Nov 30 04:02:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	130001	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	209841	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89743	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70664	51.568	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.140%
35) Dibromofluoromethane	5.391	113	63414	49.253	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.500%
50) Toluene-d8	8.653	98	241834	49.957	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.920%
62) 4-Bromofluorobenzene	11.085	95	86758	47.374	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.740%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	2789	3.663	ug/l	99
52) Toluene	8.720	92	1003	0.277	ug/l	83
70) Styrene	10.658	104	1916	0.437	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
Data File : VX025374.D
Acq On : 29 Nov 2021 16:58
Operator : JC/MD
Sample : M4837-11 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20931

Quant Time: Nov 30 04:02:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

