

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
 Data File : VX025375.D
 Acq On : 29 Nov 2021 17:21
 Operator : JC/MD
 Sample : M4837-12 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20932

Quant Time: Nov 30 04:02:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	208419	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189169	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90011	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69818	51.712	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.420%	
35) Dibromofluoromethane	5.391	113	63388	49.569	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.140%	
50) Toluene-d8	8.647	98	240395	49.998	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.000%	
62) 4-Bromofluorobenzene	11.079	95	87049	47.857	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.720%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	4367	5.821	ug/l	99
18) Methyl Acetate	2.709	43	746	0.311	ug/l #	78
40) Benzene	6.037	78	2725	0.495	ug/l #	32
52) Toluene	8.720	92	1700	0.473	ug/l	94
68) m/p-Xylenes	10.305	106	710	0.261	ug/l	93
70) Styrene	10.659	104	3806	0.865	ug/l #	87
95) Naphthalene	13.780	128	4330	0.340	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
Data File : VX025375.D
Acq On : 29 Nov 2021 17:21
Operator : JC/MD
Sample : M4837-12 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20932

Quant Time: Nov 30 04:02:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

