

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
 Data File : VX025376.D
 Acq On : 29 Nov 2021 17:45
 Operator : JC/MD
 Sample : M4837-13 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20933

Quant Time: Nov 30 04:03:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	127838	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	205338	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179802	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83080	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69378	51.487	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	102.980%	
35) Dibromofluoromethane	5.391	113	61957	49.177	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.360%	
50) Toluene-d8	8.653	98	232128	49.003	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	98.000%	
62) 4-Bromofluorobenzene	11.085	95	81014	45.207	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	90.420%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1943	2.595	ug/l	95
40) Benzene	6.043	78	7203	1.327	ug/l #	72
52) Toluene	8.720	92	3980	1.123	ug/l	98
68) m/p-Xylenes	10.305	106	1450	0.561	ug/l	90
69) o-Xylene	10.640	106	743	0.294	ug/l	96
70) Styrene	10.659	104	5154	1.232	ug/l #	89
84) 1,2,4-Trimethylbenzene	11.762	105	1048	0.213	ug/l	81
95) Naphthalene	13.780	128	6057	0.711	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
Data File : VX025376.D
Acq On : 29 Nov 2021 17:45
Operator : JC/MD
Sample : M4837-13 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20933

Quant Time: Nov 30 04:03:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

