

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
 Data File : VX025379.D
 Acq On : 29 Nov 2021 18:55
 Operator : JC/MD
 Sample : M4837-09 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20929

Quant Time: Nov 30 04:04:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126096	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	205652	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186966	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89848	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69559	52.334	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.660%	
35) Dibromofluoromethane	5.385	113	61929	49.079	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.160%	
50) Toluene-d8	8.653	98	237805	50.125	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.240%	
62) 4-Bromofluorobenzene	11.085	95	85026	47.374	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.740%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2208	2.990	ug/l	98
52) Toluene	8.720	92	57290	16.145	ug/l	98
64) Tetrachloroethene	9.268	164	1037	0.673	ug/l #	67
67) Ethyl Benzene	10.195	91	8972	1.363	ug/l	92
68) m/p-Xylenes	10.299	106	16123	5.995	ug/l	95
69) o-Xylene	10.646	106	8025	3.049	ug/l	92
70) Styrene	10.658	104	1884	0.433	ug/l #	47

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
Data File : VX025379.D
Acq On : 29 Nov 2021 18:55
Operator : JC/MD
Sample : M4837-09 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20929

Quant Time: Nov 30 04:04:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

