

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
 Data File : VX025380.D
 Acq On : 29 Nov 2021 19:18
 Operator : JC/MD
 Sample : M4837-19 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20936

Quant Time: Nov 30 04:04:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	123063	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	200482	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	183690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87112	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67777	52.250	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.500%
35) Dibromofluoromethane	5.391	113	61211	49.761	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.520%
50) Toluene-d8	8.653	98	230947	49.935	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.860%
62) 4-Bromofluorobenzene	11.079	95	83548	47.750	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.500%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	144767	200.846	ug/l	94
25) 2-Butanone	4.568	43	1609	1.544	ug/l #	83
95) Naphthalene	13.780	128	18331	2.764	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
Data File : VX025380.D
Acq On : 29 Nov 2021 19:18
Operator : JC/MD
Sample : M4837-19 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20936

Quant Time: Nov 30 04:04:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

